

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2251 SAS No.: Q2251 SDG No.: Q2251

Instrument ID: MSVOA_N Calibration Date/Time: 06/11/2025 11:32

Lab File ID: VN086940.D Init. Calib. Date(s): 06/06/2025 06/06/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:44 14:49

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.644	0.555	0.1	-13.82	20
Vinyl Chloride	0.664	0.701		5.57	20
Bromomethane	0.372	0.312		-16.13	20
Chloroethane	0.429	0.457		6.53	20
Trichlorofluoromethane	0.868	0.930		7.14	20
1,1,2-Trichlorotrifluoroethane	0.545	0.582		6.79	20
1,1-Dichloroethene	0.557	0.593		6.46	20
Acetone	0.355	0.384		8.17	20
Carbon Disulfide	1.539	1.505		-2.21	20
Methyl tert-butyl Ether	2.015	2.172		7.79	20
Methylene Chloride	0.665	0.673		1.2	20
trans-1,2-Dichloroethene	0.619	0.633		2.26	20
1,1-Dichloroethane	1.120	1.178	0.1	5.18	20
2-Butanone	0.577	0.568		-1.56	20
Carbon Tetrachloride	0.433	0.472		9.01	20
cis-1,2-Dichloroethene	0.740	0.793		7.16	20
Chloroform	1.118	1.176		5.19	20
1,1,1-Trichloroethane	0.951	0.976		2.63	20
Methylcyclohexane	0.605	0.576		-4.79	20
Benzene	1.444	1.554		7.55	20
1,2-Dichloroethane	0.438	0.472		7.76	20
Trichloroethene	0.342	0.382		11.7	20
1,2-Dichloropropane	0.351	0.384		9.4	20
Bromodichloromethane	0.480	0.532		10.83	20
4-Methyl-2-Pentanone	0.543	0.590		8.66	20
Toluene	0.882	0.978		10.88	20
t-1,3-Dichloropropene	0.537	0.609		13.41	20
cis-1,3-Dichloropropene	0.574	0.661		15.16	20
1,1,2-Trichloroethane	0.340	0.383		12.65	20
2-Hexanone	0.350	0.395		12.86	20
Dibromochloromethane	0.354	0.414		16.95	20
Tetrachloroethene	0.316	0.337		6.65	20
Chlorobenzene	1.103	1.222	0.3	10.79	20
Ethyl Benzene	1.900	2.059		8.37	20
m/p-Xylenes	0.727	0.802		10.32	20
o-Xylene	0.696	0.782		12.36	20
Styrene	1.192	1.358		13.93	20
Bromoform	0.263	0.319	0.1	21.29	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q2251 SAS No.: Q2251 SDG No.: Q2251

Instrument ID: MSVOA_N Calibration Date/Time: 06/11/2025 11:32

Lab File ID: VN086940.D Init. Calib. Date(s): 06/06/2025 06/06/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:44 14:49

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	3.643	3.898		7	20
1,1,2,2-Tetrachloroethane	1.234	1.402	0.3	13.61	20
1,3-Dichlorobenzene	1.644	1.843		12.1	20
1,4-Dichlorobenzene	1.676	1.887		12.59	20
1,2-Dichlorobenzene	1.579	1.761		11.53	20
1,2-Dichloroethane-d4	0.669	0.622		-7.03	20
Dibromofluoromethane	0.296	0.309		4.39	20
Toluene-d8	1.173	1.181		0.68	20
4-Bromofluorobenzene	0.436	0.440		0.92	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.