

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
 Lab Code: CHEM Case No.: Q2251 SAS No.: Q2251 SDG NO.: Q2251
 Lab File ID: VN086913.D Date Analyzed: 06/10/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:13
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	239452	8.23	422119	9.11	362285	11.87
UPPER LIMIT	478904	8.73	844238	9.606	724570	12.365
LOWER LIMIT	119726	7.73	211060	8.606	181143	11.365
EPA SAMPLE NO.						
BP-VPB-182-TB-20250603	351776	8.23	656962	9.11	575035	11.87
BP-VPB-182-GW-740-742	345182	8.23	642235	9.11	564142	11.87
BP-VPB-182-GW-760-762	337846	8.23	630375	9.11	558009	11.87
BP-VPB-182-EB-20250604	207035	8.24	385111	9.11	340856	11.87
VN0610WBL01	241251	8.23	440901	9.11	386425	11.87
VN0610WBS01	249843	8.23	443174	9.11	389297	11.87
VN0610WBSD01	231054	8.23	413772	9.11	362257	11.87

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.