

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q2251 SAS No.: Q2251 SDG NO.: Q2251
Lab File ID: VY022652.D Date Analyzed: 06/11/2025
Instrument ID: MSVOA_Y Time Analyzed: 09:59
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	157578	7.71	265316	8.62	226799	11.41
UPPER LIMIT	315156	8.207	530632	9.116	453598	11.914
LOWER LIMIT	78789	7.207	132658	8.116	113400	10.914
EPA SAMPLE NO.						
BP-VPB-182-GW-780-782	210051	7.71	390122	8.62	315222	11.41
VY0611SBL01	167932	7.71	308222	8.62	245642	11.41
VY0611SBS01	154377	7.71	263261	8.62	218934	11.41
VY0611SBSD01	154293	7.71	261377	8.62	216375	11.41

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.