

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q2251 SAS No.: Q2251 SDG NO.: Q2251
Lab File ID: VN086913.D Date Analyzed: 06/10/2025
Instrument ID: MSVOA_N Time Analyzed: 09:13
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	170987	13.788				
UPPER LIMIT	341974	14.288				
LOWER LIMIT	85493.5	13.288				
EPA SAMPLE NO.						
BP-VPB-182-TB-20250603	273719	13.79				
BP-VPB-182-GW-740-742	270892	13.79				
BP-VPB-182-GW-760-762	267650	13.79				
BP-VPB-182-EB-20250604	162809	13.79				
VN0610WBL01	183380	13.79				
VN0610WBS01	189620	13.79				
VN0610WBSD01	176612	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.