

DATA PACKAGE GENERAL CHEMISTRY

PROJECT NAME : NWIRP BETHPAGE 112G08005-WE13

**TETRA TECH NUS, INC.
661 Andersen Drive
Suite 200
Pittsburgh, PA - 15220-2745
Phone No: 412-921-7090**

**ORDER ID : Q2253
ATTENTION : Ernie Wu**



Laboratory Certification ID # 20012



1) GENERAL CHEMISTRY DATA	2
2) Signature Page	3
3) Case Narrative	4
4) Qualifier Page	5
5) Conformance/Non Conformance	6
6) QA Checklist	7
7) Chronicle	8
8) Sample Data	9
8.1) RW8-SP303-20250605	10
9) QC Data Summary For Genchem	11
9.1) Preparation Blank Summary	12
9.2) Duplicate Sample Summary	13
9.3) Laboratory Control Sample Summary	15
10) GENCHEM RAW DATA	17
10.1) GENCHEM RAW DATA - ANALYTICAL	18
10.1.1) LB136041	18
10.1.2) LB136054	20
11) Analytical Runlogs	22
12) Standard Prep Logs	24
13) Shipping Document	26
13.1) Chain Of Custody	27
13.2) Lab Certificate	28

1
2
3
4
5
6
7
8
9
10
11
12
13

Cover Page

Order ID : Q2253

Project ID : NWIRP Bethpage 112G08005-WE13

Client : Tetra Tech NUS, Inc.

Lab Sample Number

Q2253-01
Q2253-02

Client Sample Number

RW8-SP100-20250605
RW8-SP303-20250605

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature :

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 4:11 pm, Jun 17, 2025

Date: 6/11/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Tetra Tech NUS, Inc.

Project Name: NWIRP Bethpage 112G08005-WE13

Project Manager: Ernie Wu

Order ID # Q2253

Test Name: TDS,TSS

A. Number of Samples and Date of Receipt:

2 Water samples were received on 06/05/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Metals Group4, SVOC-SIMGroup1, TDS and TSS. This data package contains results for TDS,TSS.

C. Analytical Techniques:

The analysis of TDS was based on method SM2540 C and The analysis of TSS was based on method SM2540 D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

E. Additional Comments:

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is).

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 4:11 pm, Jun 17, 2025

DATA REPORTING QUALIFIERS- INORGANIC

For reporting results, the following “ Results Qualifiers” are used:

J	Indicates the reported value was obtained from a reading that was less than the Contract Required Detection Limit (CRDL), but greater than or equal to the Instrument Detection Limit (IDL).
U	Indicates the analyte was analyzed for, but not detected.
ND	Indicates the analyte was analyzed for, but not detected
E	Indicates the reported value is estimated because of the presence of interference
M	Indicates Duplicate injection precision not met.
N	Indicates the spiked sample recovery is not within control limits.
S	Indicates the reported value was determined by the Method of Standard Addition (MSA).
*	Indicates that the duplicate analysis is not within control limits.
+	Indicates the correlation coefficient for the MSA is less than 0.995.
D	Indicates the reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.
M	Method qualifiers “P” for ICP instrument “PM” for ICP when Microwave Digestion is used “CV” for Manual Cold Vapor AA “AV” for automated Cold Vapor AA “CA” for MIDI-Distillation Spectrophotometric “AS” for Semi -Automated Spectrophotometric “C” for Manual Spectrophotometric “T” for Titrimetric “NR” for analyte not required to be analyzed
OR	Indicates the analyte’s concentration exceeds the calibrated range of the instrument for that specific analysis.
Q	Indicates the LCS did not meet the control limits requirements
H	Sample Analysis Out Of Hold Time

GENERAL CHEMISTRY CONFORMANCE/NON-CONFORMANCE SUMMARY

ORDER ID: Q2253

MATRIX: Water

METHOD: SM2540 C,SM2540 D

	NA	NO	YES
1. Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
2. Matrix Spike Duplicate Recoveries Met Criteria			✓
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
The Blank Spike met requirements for all samples.			
3. Sample Duplicate Analysis Met QC Criteria			✓
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
4. Digestion Holding Time Met			✓
If not met, list number of days exceeded for each sample:			

ADDITIONAL COMMENTS: The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is).

REVIEWED

QA REVIEW

By Sohil Jodhani, QA/QC Director at 11:34 am, Jun 17, 2025

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q2253

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 06/11/2025

LAB CHRONICLE

OrderID:	Q2253	OrderDate:	6/5/2025 4:32:00 PM
Client:	Tetra Tech NUS, Inc.	Project:	NWIRP Bethpage 112G08005-WE13
Contact:	Ernie Wu	Location:	D21

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2253-02	RW8-SP303-2025060 5	WATER			06/05/25 13:13			06/05/25
			TDS	SM2540 C			06/06/25 12:30	
			TSS	SM2540 D			06/09/25 13:00	



SAMPLE DATA

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	06/05/25 13:13
Project:	NWIRP Bethpage 112G08005-WE13	Date Received:	06/05/25
Client Sample ID:	RW8-SP303-20250605	SDG No.:	Q2253
Lab Sample ID:	Q2253-02	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
TDS	321		1	1.00	10.0	10.0	mg/L		06/06/25 12:30	SM 2540 C-15
TSS	4.00	U	1	1.00	4.00	4.00	mg/L		06/09/25 13:00	SM 2540 D-15

Comments: _____

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

QC RESULT SUMMARY

Preparation Blank Summary

Client: Tetra Tech NUS, Inc.

SDG No.: Q2253

Project: NWIRP Bethpage 112G08005-WE13

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: LB136041BL							
TDS	mg/L	< 5.0000	5.0000	U	1.0	10	06/06/2025
Sample ID: LB136054BL							
TSS	mg/L	1	2.0000	J	1	4	06/09/2025

Duplicate Sample Summary

Client:	Tetra Tech NUS, Inc.	SDG No.:	Q2253
Project:	NWIRP Bethpage 112G08005-WE13	Sample ID:	Q2229-02
Client ID:	COMPDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
TSS	mg/L	+/-5	121		121		1	0		06/09/2025

Duplicate Sample Summary

Client:	Tetra Tech NUS, Inc.	SDG No.:	Q2253
Project:	NWIRP Bethpage 112G08005-WE13	Sample ID:	Q2231-05
Client ID:	MW-17-20250604DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
TDS	mg/L	+/-5	296		302		1	2.01		06/06/2025

Laboratory Control Sample Summary

Client:	Tetra Tech NUS, Inc.	SDG No.:	Q2253
Project:	NWIRP Bethpage 112G08005-WE13	Run No.:	LB136041

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136041BS							
TDS	mg/L	100	95.0		95	1	90-110	06/06/2025

Laboratory Control Sample Summary

Client:	Tetra Tech NUS, Inc.	SDG No.:	Q2253
Project:	NWIRP Bethpage 112G08005-WE13	Run No.:	LB136054

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136054BS							
TSS	mg/L	550	533		97	1	90-110	06/09/2025



RAW DATA

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13

WORKLIST(Hardcopy Internal Chain)

136041

WorkList Name : tds q2253

WorkList ID : 190001

Department : Wet-Chemistry

Date : 06-06-2025 11:57:55

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2231-01	MW-10D-20250604	Water	TDS	Cool 4 deg C	PARS02	N41	06/04/2025	SM2540 C
Q2231-03	MW-15-20250604	Water	TDS	Cool 4 deg C	PARS02	N41	06/04/2025	SM2540 C
Q2231-04	MW-16D-20250604	Water	TDS	Cool 4 deg C	PARS02	N41	06/04/2025	SM2540 C
Q2231-05	MW-17-20250604	Water	TDS	Cool 4 deg C	PARS02	N41	06/04/2025	SM2540 C
Q2234-01	MW-17B-55-060425	Water	TDS	Cool 4 deg C	JACO05	N31	06/04/2025	SM2540 C
Q2234-02	Q2234-01MS	Water	TDS	Cool 4 deg C	JACO05	N31	06/04/2025	SM2540 C
Q2234-03	Q2234-01MSD	Water	TDS	Cool 4 deg C	JACO05	N31	06/04/2025	SM2540 C
Q2234-05	MW-18B-56-060425	Water	TDS	Cool 4 deg C	JACO05	N31	06/04/2025	SM2540 C
Q2234-06	MW-18B-56-060425-FD	Water	TDS	Cool 4 deg C	JACO05	N31	06/04/2025	SM2540 C
Q2234-07	MW-19B-72-060425	Water	TDS	Cool 4 deg C	JACO05	N31	06/04/2025	SM2540 C
Q2250-05	EB02-060525	Water	TDS	Cool 4 deg C	JACO05	N31	06/04/2025	SM2540 C
Q2253-02	RW8-SO303-20250605	Water	TDS	Cool 4 deg C	TETR06	D21	06/05/2025	SM2540 C

Date/Time 06/06/25 12:30

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 06/06/25

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: Iwona

ANALYST: jignesh

Date: 06/06/2025

Run Number: LB136054

BalanceID: WC SC-6

OvenID: WC OVEN-1

FilterID: 17416528

ThermometerID: WET OVEN#1

TEMP1 IN: 104 °C 06/06/2025 14:00 TEMP1 OUT: 104 °C 06/06/2025 15:00
 TEMP2 IN: 103 °C 06/06/2025 15:30 TEMP2 OUT: 103 °C 06/06/2025 16:30
 TEMP3 IN: 104 °C 06/09/2025 13:00 TEMP3 OUT: 103 °C 06/09/2025 14:30
 TEMP4 IN: 104 °C 06/09/2025 15:00 TEMP4 OUT: 104 °C 06/09/2025 16:30

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB136054BL	LB136054BL	1.3562	1.3562	100	1.3563	1.3563	1.3563	0.0001	1
2	LB136054BS	LB136054BS	1.5893	1.5893	100	1.6426	1.6426	1.6426	0.0533	533
3	Q2229-02	COMP	1.4768	1.4768	200	1.5010	1.5010	1.5010	0.0242	121
4	Q2229-02DUP	COMPDUP	1.4924	1.4924	200	1.5166	1.5166	1.5166	0.0242	121
5	Q2237-02	TW-WTS-10	1.4886	1.4886	1000	1.5091	1.5091	1.5091	0.0205	20.5
6	Q2243-01	WATER-TREATMENT-DISCHARGE	1.4729	1.4729	1000	1.4824	1.4824	1.4824	0.0095	9.5
7	Q2253-02	RW8-SO303-20250605	1.5000	1.5000	800	1.5006	1.5006	1.5006	0.0006	0.7
8	Q2264-04	EFF-WW	1.4907	1.4907	700	1.5078	1.5078	1.5078	0.0171	24.4

A = Sample Volume (ml)
 B = Final Empty Dish Weight (g)
 C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)
 D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

LB136054

WorkList Name : tss q2243

WorkList ID : 190041

Department : Wet-Chemistry

Date : 06-09-2025 10:48:23

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2229-02	B COMP	Water	TSS	Cool 4 deg C	ARAM01	N41	06/04/2025	SM2540 D
Q2237-02	E TW-WTS-10	Water	TSS	Cool 4 deg C	ENTA05	N31	06/04/2025	SM2540 D
Q2243-01	E WATER-TREATMENT-DISCHAI	Water	TSS	Cool 4 deg C	VERI01	N41	06/05/2025	SM2540 D
Q2253-02	C RW8-SO303-20250605	Water	TSS	Cool 4 deg C	TETRO6	D21	06/05/2025	SM2540 D
Q2264-04	B EFF-WW	Water	TSS	Cool 4 deg C	ARDM01	D41	06/06/2025	SM2540 D

Date/Time

06/09/25 11:30

Raw Sample Received by:

SGWOC

Raw Sample Relinquished by:

SGWOC

Date/Time

06/09/25

Raw Sample Received by:

SGWOC

Raw Sample Relinquished by:

SGWOC

Instrument ID: WC SC-3

Daily Analysis Runlog For Sequence/QC Batch ID # LB136041

Review By	jignesh	Review On	6/9/2025 10:21:18 AM
Supervise By	Iwona	Supervise On	6/9/2025 1:13:30 PM
SubDirectory	LB136041	Test	TDS
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	N/A		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	LB136041BL	LB136041BL	MB	06/06/25 12:30		jignesh	OK
2	LB136041BS	LB136041BS	LCS	06/06/25 12:30		jignesh	OK
3	Q2231-01	MW-10D-20250604	SAM	06/06/25 12:30		jignesh	OK
4	Q2231-03	MW-15-20250604	SAM	06/06/25 12:30		jignesh	OK
5	Q2231-04	MW-16D-20250604	SAM	06/06/25 12:30		jignesh	OK
6	Q2231-05	MW-17-20250604	SAM	06/06/25 12:30		jignesh	OK
7	Q2231-05DUP	MW-17-20250604DUP	DUP	06/06/25 12:30		jignesh	OK
8	Q2234-01	MW-17B-55-060425	SAM	06/06/25 12:30		jignesh	OK
9	Q2234-05	MW-18B-56-060425	SAM	06/06/25 12:30		jignesh	OK
10	Q2234-06	MW-18B-56-060425-F	SAM	06/06/25 12:30		jignesh	OK
11	Q2234-07	MW-19B-72-060425	SAM	06/06/25 12:30		jignesh	OK
12	Q2250-05	EB02-060525	SAM	06/06/25 12:30		jignesh	OK
13	Q2253-02	RW8-SO303-2025060	SAM	06/06/25 12:30		jignesh	OK

Instrument ID: WC SC-3

Daily Analysis Runlog For Sequence/QC Batch ID # LB136054

Review By	jignesh	Review On	6/9/2025 11:54:33 AM
Supervise By	Iwona	Supervise On	6/9/2025 1:08:21 PM
SubDirectory	LB136054	Test	TSS
STD. NAME	STD REF.#		
ICAL Standard	N/A		
ICV Standard	N/A		
CCV Standard	N/A		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	N/A		
Chk Standard	N/A		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	LB136054BL	LB136054BL	MB	06/09/25 13:00		jignesh	OK
2	LB136054BS	LB136054BS	LCS	06/09/25 13:00		jignesh	OK
3	Q2229-02	COMP	SAM	06/09/25 13:00		jignesh	OK
4	Q2229-02DUP	COMPDUP	DUP	06/09/25 13:00		jignesh	OK
5	Q2237-02	TW-WTS-10	SAM	06/09/25 13:00		jignesh	OK
6	Q2243-01	WATER-TREATMENT	SAM	06/09/25 13:00		jignesh	OK
7	Q2253-02	RW8-SO303-2025060	SAM	06/09/25 13:00		jignesh	OK
8	Q2264-04	EFF-WW	SAM	06/09/25 13:00		jignesh	OK

Prep Standard - Chemical Standard Summary

Order ID : Q2253

Test : TDS,TSS

Prepbatch ID :

Sequence ID/Qc Batch ID: LB136041, LB136054,

Standard ID :

Chemical ID :

1
2
3
4
5
6
7
8
9
10
11
12
13

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
				/	/	

1
2
3
4
5
6
7
8
9
10
11
12
13



SHIPPING DOCUMENTS

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13

 CHAIN OF CUSTODY RECORD		284 Sheffield Street, Mountainside, NJ 07092 (908) 789-8900 Fax: (908) 78-8922 www.chemtech.net		Chemtech Project Number: Q2253																					
CLIENT INFORMATION		PROJECT INFORMATION		BILLING INFORMATION																					
COMPANY: Tetra Tech		PROJECT NAME: NWIRP Bethpage		BILL TO: PO#																					
ADDRESS: 4433 Corporation Ln, Suite 300		PROJECT #: 112G08005-WE13 LOCATION: RW8		ADDRESS:																					
CITY: Virginia Beach STATE: VA ZIP: 23462		PROJECT MANAGER: Ernie Wu		CITY: STATE: ZIP:																					
ATTENTION: Ernie Wu		E-MAIL: ernie.wu@tetratech.com		ATTENTION: PHONE:																					
PHONE: 757-466-4901 FAX: 757-461-4148		PHONE: 757-466-4901 FAX: 757-461-4148																							
DATA TURNAROUND INFORMATION		DATA DELIVERABLE INFORMATION		ANALYSIS																					
FAX: 5 DAYS* HARD COPY: 5 DAYS* EDD 5 DAYS* * TO BE APPROVED BY CHEMTECH STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS		☐ RESULTS ONLY ☐ USEPA CLP ☐ RESULTS + QC ☐ New York State ASP "B" ☐ New Jersey REDUCED ☐ New York State ASP "A" ☐ New Jersey CLP ☐ Other _____ ☐ EDD Format _____		<table border="1" style="width:100%; border-collapse: collapse; font-size: 0.8em;"> <tr> <td style="width: 10%;">1,4-Dioxane SW846 8270</td> <td style="width: 10%;">SIM</td> <td style="width: 10%;">Iron, Total</td> <td style="width: 10%;">TSS</td> <td style="width: 10%;">TDS</td> <td style="width: 10%;"> </td> <td style="width: 10%;"> </td> <td style="width: 10%;"> </td> <td style="width: 10%;"> </td> <td style="width: 10%;"> </td> </tr> <tr> <td>1</td><td>2</td><td>3</td><td>4</td><td>5</td><td>6</td><td>7</td><td>8</td><td>9</td><td></td> </tr> </table>		1,4-Dioxane SW846 8270	SIM	Iron, Total	TSS	TDS						1	2	3	4	5	6	7	8	9	
1,4-Dioxane SW846 8270	SIM	Iron, Total	TSS	TDS																					
1	2	3	4	5	6	7	8	9																	
CHEMTECH SAMPLE ID		PROJECT SAMPLE IDENTIFICATION		SAMPLE MATRIX		SAMPLE TYPE		SAMPLE COLLECTION		# of Bottles		PRESERVATIVES		COMMENTS											
						COMP GRAB		DATE TIME				1 2 3 4 5 6 7 8 9		<-- Specify Preservatives A-HCl B-HNO3 C-H2SO4 D-NaOH E-ICE F-Other											
1.		RW8-SP100-20250605		GW		X		6/5/25 13:05		2		X X													
2.		RW8-SP303-20250605		GW		X		6/5/25 13:13		4		X X X X													
3.																									
4.																									
5.																									
6.																									
7.																									
8.																									
9.																									
10.																									
SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE PROSESSION INCLUDING COURIER DELIVERY																									
RELINQUISHED BY SAMPLER		DATE/TIME		RECEIVED BY		Conditions of bottles or coolers at receipt: ☐ Compliant ☐ Non Compliant ☐ Cooler Temp <u>2.1C</u> MeOH extraction requires an additional 4oz. Jar for percent solid ☐ Ice in Cooler?: _____																			
1.		6/5/25 15:00		1.		Comments: Page _____ of _____																			
RELINQUISHED BY		DATE/TIME		RECEIVED BY																					
2.				2.																					
RELINQUISHED BY		DATE/TIME		RECEIVED FOR LAB BY		SHIPPED VIA: CLIENT: ☐ Hand Delivered ☐ Overnight CHEMTECH: ☐ Picked Up ☐ Overnight				Shipment Complete ☐ YES ☐ NO															
3.		6-5-25 19:20		3.																					

WHITE - CHEMTECH COPY FOR RETURN TO CLIENT

YELLOW - CHEMTECH COPY

PINK - SAMPLER COPY

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488