

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** Q2253
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q2253 **SAS No.:** Q2253
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	3770	4000	94	90 - 110	P	06/10/2025	14:37	LB136097

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** Q2253

Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q2253 **SAS No.:** Q2253

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	108	100	108	80 - 120	P	06/10/2025	14:41	LB136097

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** Q2253

Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q2253 **SAS No.:** Q2253

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	4860	5000	97	90 - 110	P	06/10/2025	15:30	LB136097
CCV02	Iron	4930	5000	99	90 - 110	P	06/10/2025	16:28	LB136097
CCV03	Iron	5390	5000	108	90 - 110	P	06/10/2025	17:36	LB136097
CCV04	Iron	5420	5000	108	90 - 110	P	06/10/2025	18:28	LB136097
CCV05	Iron	5340	5000	107	90 - 110	P	06/10/2025	19:18	LB136097
CCV06	Iron	5150	5000	103	90 - 110	P	06/10/2025	20:09	LB136097
CCV07	Iron	5190	5000	104	90 - 110	P	06/10/2025	21:04	LB136097
CCV08	Iron	5140	5000	103	90 - 110	P	06/10/2025	22:04	LB136097
CCV09	Iron	5160	5000	103	90 - 110	P	06/10/2025	22:56	LB136097
CCV10	Iron	5200	5000	104	90 - 110	P	06/10/2025	23:57	LB136097
CCV11	Iron	5060	5000	101	90 - 110	P	06/11/2025	01:35	LB136097
CCV12	Iron	4930	5000	99	90 - 110	P	06/11/2025	02:21	LB136097