

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tetra Tech NUS, Inc. **SDG No.:** Q2253
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q2253 **SAS No.:** Q2253
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Iron	97000	101000	96	85600	116500	06/10/2025	14:54	LB136097
ICSAB01	Iron	99000	99300	100	84400	114500	06/10/2025	15:18	LB136097