

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: Q2253

SDG NO.: Q2253

Lab File ID: BN037188.D

DFTPP Injection Date: 06/09/2025

Instrument ID: BNA_N

DFTPP Injection Time: 10:15

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	74
68	Less than 2.0% of mass 69	0.4 (0.7) 1
69	Mass 69 relative abundance	59.6
70	Less than 2.0% of mass 69	0.4 (0.6) 1
127	10.0 - 80.0% of mass 198	53
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
275	10.0 - 60.0% of mass 198	24.1
365	Greater than 1% of mass 198	4.4
441	Present, but less than mass 443	8.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	10.5 (18.5) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN037189.D	06/09/2025	10:54
PB168336BL	PB168336BL	BN037190.D	06/09/2025	11:30
MW-11A-13.5-060525MS	Q2250-02MS	BN037192.D	06/09/2025	14:33
MW-11A-13.5-060525MSD	Q2250-03MSD	BN037193.D	06/09/2025	15:47
RW8-SP100-20250605	Q2253-01	BN037197.D	06/09/2025	18:15
RW8-SP303-20250605	Q2253-02	BN037198.D	06/09/2025	18:51
PB168336BS	PB168336BS	BN037201.D	06/09/2025	20:40
SSTDCCC0.4EC	SSTDCCC0.4	BN037202.D	06/09/2025	21:16