

Surrogate Summary

SDG No.: Q2254

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2254-01	BP-VPB-182-GW-810-812	1,2-Dichloroethane-d4	50	42.8	86	81	118
		Dibromofluoromethane	50	48.3	97	80	119
		Toluene-d8	50	50.0	100	89	112
		4-Bromofluorobenzene	50	48.0	96	85	114
VN0609WBL01	VN0609WBL01	1,2-Dichloroethane-d4	50	42.8	86	81	118
		Dibromofluoromethane	50	48.1	96	80	119
		Toluene-d8	50	50.7	101	89	112
		4-Bromofluorobenzene	50	48.9	98	85	114
VN0609WBS01	VN0609WBS01	1,2-Dichloroethane-d4	50	56.6	113	81	118
		Dibromofluoromethane	50	58.6	117	80	119
		Toluene-d8	50	55.2	110	89	112
		4-Bromofluorobenzene	50	55.7	111	85	114
VN0609WBSD0	VN0609WBSD01	1,2-Dichloroethane-d4	50	44.5	89	81	118
		Dibromofluoromethane	50	50.5	101	80	119
		Toluene-d8	50	48.2	96	89	112
		4-Bromofluorobenzene	50	47.7	95	85	114