

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2254

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Datafile : VN086902.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0609WBSD01	Chloromethane	20	15.5	ug/L	78	7		50	139	20
	Vinyl chloride	20	18.2	ug/L	91	1		58	137	20
	Bromomethane	20	17.0	ug/L	85	20		53	141	20
	Chloroethane	20	18.2	ug/L	91	7		60	138	20
	Trichlorofluoromethane	20	19.1	ug/L	96	0		65	141	20
	1,1,2-Trichlorotrifluoroethane	20	18.2	ug/L	91	5		70	136	20
	1,1-Dichloroethene	20	19.5	ug/L	98	3		71	131	20
	Acetone	100	76.9	ug/L	77	24	*	39	160	20
	Carbon disulfide	20	17.5	ug/L	88	2		64	133	20
	Methyl tert-butyl Ether	20	19.2	ug/L	96	18		71	124	20
	Methylene Chloride	20	17.9	ug/L	90	13		74	124	20
	trans-1,2-Dichloroethene	20	18.7	ug/L	94	5		75	124	20
	1,1-Dichloroethane	20	18.4	ug/L	92	7		77	125	20
	2-Butanone	100	82.3	ug/L	82	20		56	143	20
	Carbon Tetrachloride	20	19.4	ug/L	97	0		72	136	20
	cis-1,2-Dichloroethene	20	18.9	ug/L	95	11		78	123	20
	Chloroform	20	18.4	ug/L	92	8		79	124	20
	1,1,1-Trichloroethane	20	18.6	ug/L	93	4		74	131	20
	Methylcyclohexane	20	16.3	ug/L	81	6		72	132	20
	Benzene	20	19.0	ug/L	95	6		79	120	20
	1,2-Dichloroethane	20	18.8	ug/L	94	13		73	128	20
	Trichloroethene	20	20.4	ug/L	102	2		79	123	20
	1,2-Dichloropropane	20	18.6	ug/L	93	11		78	122	20
	Bromodichloromethane	20	19.2	ug/L	96	10		79	125	20
	4-Methyl-2-Pentanone	100	89.3	ug/L	89	21	*	67	130	20
	Toluene	20	19.8	ug/L	99	11		80	121	20
	t-1,3-Dichloropropene	20	19.8	ug/L	99	13		73	127	20
	cis-1,3-Dichloropropene	20	19.6	ug/L	98	12		75	124	20
	1,1,2-Trichloroethane	20	19.7	ug/L	99	14		80	119	20
	2-Hexanone	100	86.5	ug/L	87	23	*	57	139	20
	Dibromochloromethane	20	20.4	ug/L	102	11		74	126	20
	Tetrachloroethene	20	19.2	ug/L	96	2		74	129	20
	Chlorobenzene	20	20.3	ug/L	102	2		82	118	20
	Ethyl Benzene	20	19.5	ug/L	98	1		79	121	20
	m/p-Xylenes	40	39.7	ug/L	99	3		80	121	20
	o-Xylene	20	20.1	ug/L	101	3		78	122	20
	Styrene	20	20.1	ug/L	101	5		78	123	20
	Bromoform	20	21.5	ug/L	108	10		66	130	20
	Isopropylbenzene	20	20.0	ug/L	100	0		72	131	20
	1,1,2,2-Tetrachloroethane	20	21.5	ug/L	108	8		71	121	20
	1,3-Dichlorobenzene	20	20.6	ug/L	103	3		80	119	20
	1,4-Dichlorobenzene	20	20.7	ug/L	104	2		79	118	20
	1,2-Dichlorobenzene	20	21.1	ug/L	106	1		80	119	20