

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06  
Lab Code: CHEM Case No.: Q2254 SAS No.: Q2254 SDG NO.: Q2254  
Lab File ID: VN086888.D Date Analyzed: 06/09/2025  
Instrument ID: MSVOA\_N Time Analyzed: 08:37  
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

|                       | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #   |
|-----------------------|---------------|-------|---------------|-------|---------------|--------|
| 12 HOUR STD           | 261450        | 8.23  | 451735        | 9.11  | 380922        | 11.87  |
| UPPER LIMIT           | 522900        | 8.729 | 903470        | 9.606 | 761844        | 12.365 |
| LOWER LIMIT           | 130725        | 7.729 | 225868        | 8.606 | 190461        | 11.365 |
| EPA SAMPLE NO.        |               |       |               |       |               |        |
| BP-VPB-182-GW-810-812 | 313915        | 8.23  | 549826        | 9.11  | 462153        | 11.87  |
| VN0609WBL01           | 390256        | 8.23  | 684606        | 9.11  | 591728        | 11.87  |
| VN0609WBS01           | 266986        | 8.23  | 481053        | 9.11  | 422488        | 11.87  |
| VN0609WBSD01          | 261604        | 8.23  | 462540        | 9.11  | 392287        | 11.87  |

IS1 = Pentafluorobenzene  
IS2 = 1,4-Difluorobenzene  
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
AREA LOWER LIMIT = -50% of internal standard area  
RT UPPER LIMIT = +0.50 minutes of internal standard RT  
RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
\* Values outside of QC limits.