

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CHEM02

Lab Code: CHEM **Case No.:** Q2258 **SAS No.:** Q2258 **SDG NO.:** Q2258

Instrument ID: ECD_D **Date(s) Analyzed:** _____

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	6.24	6.14	6.34	29917300
		2	6.24	6.14	6.34	28630300
		3	6.24	6.14	6.34	27430700
		4	6.24	6.14	6.34	27784800
		5	6.24	6.14	6.34	28514500
Toxaphene	1000	1	6.44	6.34	6.54	40532500
		2	6.44	6.34	6.54	39836800
		3	6.44	6.34	6.54	39009200
		4	6.44	6.34	6.54	39639000
		5	6.44	6.34	6.54	40793000
Toxaphene	2000	1	7.06	6.96	7.16	104833000
		2	7.06	6.96	7.16	104771000
		3	7.06	6.96	7.16	106716000
		4	7.06	6.96	7.16	109901000
		5	7.06	6.96	7.16	114900000
Toxaphene	4000	1	7.15	7.05	7.25	77360800
		2	7.15	7.05	7.25	77708800
		3	7.15	7.05	7.25	78817000
		4	7.15	7.05	7.25	81932300
		5	7.15	7.05	7.25	85218200
Toxaphene	8000	1	7.93	7.83	8.03	58009400
		2	7.93	7.83	8.03	57341500
		3	7.93	7.83	8.03	57649700
		4	7.93	7.83	8.03	59089300
		5	7.93	7.83	8.03	61113800

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CHEM02

Lab Code: CHEM **Case No.:** Q2258 **SAS No.:** Q2258 **SDG NO.:** Q2258

Instrument ID: ECD_D **Date(s) Analyzed:** _____

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.15	5.05	5.25	127613000
		2	5.15	5.05	5.25	122562000
		3	5.15	5.05	5.25	113529000
		4	5.15	5.05	5.25	108991000
		5	5.15	5.05	5.25	102145000
Toxaphene	1000	1	6.12	6.02	6.22	141537000
		2	6.12	6.02	6.22	132514000
		3	6.12	6.02	6.22	126662000
		4	6.12	6.02	6.22	115690000
		5	6.12	6.02	6.22	112359000
Toxaphene	2000	1	6.76	6.66	6.86	422706000
		2	6.76	6.66	6.86	418318000
		3	6.76	6.66	6.86	399801000
		4	6.76	6.66	6.86	399524000
		5	6.76	6.66	6.86	391710000
Toxaphene	4000	1	7.20	7.10	7.30	300598000
		2	7.20	7.10	7.30	290563000
		3	7.20	7.10	7.30	274687000
		4	7.20	7.10	7.30	263451000
		5	7.20	7.10	7.30	252842000
Toxaphene	8000	1	7.33	7.23	7.43	213328000
		2	7.33	7.23	7.43	217701000
		3	7.33	7.23	7.43	211522000
		4	7.33	7.23	7.43	209861000
		5	7.33	7.23	7.43	208808000