

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2258 SAS No.: Q2258 SDG NO.: Q2258

EPA Sample No.: SSTD020226 Date Analyzed: 06/19/2025

Lab File ID: BN037342.D Time Analyzed: 12:19

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	69812	7.564	219058	10.34	169195	14.22
UPPER LIMIT	139624	8.064	438116	10.84	338390	14.716
LOWER LIMIT	34906	7.064	109529	9.84	84597.5	13.716
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	98516	7.56	308663	10.34	214565	14.22

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.