

Surrogate Summary

SDG No.: **Q2258**

Client: **Chemtech Consulting Group**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2258-01	STORAGE-BLANK-SOIL-REF-6	1,2-Dichloroethane-d4	50	48.2	96	74	125
		Dibromofluoromethane	50	49.2	98	75	124
		Toluene-d8	50	50.3	101	86	113
		4-Bromofluorobenzene	50	50.2	100	77	121
Q2258-02	STORAGE-BLANK-WATER-REF-5	1,2-Dichloroethane-d4	50	48.6	97	74	125
		Dibromofluoromethane	50	49.3	99	75	124
		Toluene-d8	50	50.6	101	86	113
		4-Bromofluorobenzene	50	50.2	100	77	121
Q2258-03	STORAGE-BLANK-WATER-REF-4	1,2-Dichloroethane-d4	50	50.4	101	74	125
		Dibromofluoromethane	50	49.2	98	75	124
		Toluene-d8	50	50.2	100	86	113
		4-Bromofluorobenzene	50	49.8	100	77	121
Q2258-04	STORAGE-BLANK-SAMPLE-MANAC	1,2-Dichloroethane-d4	50	47.6	95	74	125
		Dibromofluoromethane	50	48.4	97	75	124
		Toluene-d8	50	50.5	101	86	113
		4-Bromofluorobenzene	50	49.5	99	77	121
VX0620WBL02	VX0620WBL02	1,2-Dichloroethane-d4	50	47.0	94	74	125
		Dibromofluoromethane	50	48.8	98	75	124
		Toluene-d8	50	49.8	100	86	113
		4-Bromofluorobenzene	50	49.7	99	77	121
VX0620WBS02	VX0620WBS02	1,2-Dichloroethane-d4	50	48.0	96	74	125
		Dibromofluoromethane	50	49.3	99	75	124
		Toluene-d8	50	50.4	101	86	113
		4-Bromofluorobenzene	50	52.2	104	77	121
VX0620WBSD0	VX0620WBSD02	1,2-Dichloroethane-d4	50	48.2	96	74	125
		Dibromofluoromethane	50	49.8	100	75	124
		Toluene-d8	50	50.6	101	86	113
		4-Bromofluorobenzene	50	52.7	105	77	121