

Initial and Continuing Calibration Verification

Client: Nobis Group

SDG No.: Q2259

Project: Raymark Superfund Site

RunNo.: LB136063

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Cyanide	mg/L	0.095	0.099	96	90-110	06/09/2025
Sample ID: CCV1 Cyanide	mg/L	0.24	0.25	96	90-110	06/09/2025
Sample ID: CCV2 Cyanide	mg/L	0.23	0.25	92	90-110	06/09/2025
Sample ID: CCV3 Cyanide	mg/L	0.24	0.25	96	90-110	06/09/2025