

### Laboratory Control Sample Summary

**Client:** Nobis Group

**SDG No.:** Q2259

**Project:** Raymark Superfund Site

**Run No.:** LB136063

| Analyte   | Units      | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|-----------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID | PB168361BS |            |        |                 |            |                 |                     |               |
| Cyanide   | mg/Kg      | 5          | 4.70   |                 | 94         | 1               | 85-115              | 06/09/2025    |