

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Nobis Group **SDG No.:** Q2259
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q2259 **SAS No.:** Q2259
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV30	Mercury	3.95	4.0	99	90 - 110	CV	06/09/2025	11:34	LB136058

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Nobis Group **SDG No.:** Q2259
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q2259 **SAS No.:** Q2259
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV13	Mercury	4.79	5.0	96	90 - 110	CV	06/09/2025	11:44	LB136058
CCV14	Mercury	4.63	5.0	93	90 - 110	CV	06/09/2025	12:28	LB136058
CCV15	Mercury	5.22	5.0	104	90 - 110	CV	06/09/2025	12:57	LB136058
CCV16	Mercury	4.52	5.0	90	90 - 110	CV	06/09/2025	13:51	LB136058
CCV17	Mercury	4.68	5.0	94	90 - 110	CV	06/09/2025	14:19	LB136058
CCV18	Mercury	4.83	5.0	97	90 - 110	CV	06/09/2025	14:30	LB136058



- 2a -

Client:	<u>Nobis Group</u>	SDG No.:	<u>Q2259</u>				
Contract:	<u>NOBI03</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q2259</u>	SAS No.:	<u>Q2259</u>
Initial Calibration Source:	<u>EPA</u>						
Continuing Calibration Source:	Inorganic Ventures						

Sample ID	Analyte	Result	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
		ug/L							
ICV01	Aluminum	7520	8000	94	90 - 110	P	06/10/2025	14:37	LB136097
	Antimony	4020	4000	100	90 - 110	P	06/10/2025	14:37	LB136097
	Arsenic	3850	4000	96	90 - 110	P	06/10/2025	14:37	LB136097
	Barium	7450	8000	93	90 - 110	P	06/10/2025	14:37	LB136097
	Beryllium	198	200	99	90 - 110	P	06/10/2025	14:37	LB136097
	Cadmium	1920	2000	96	90 - 110	P	06/10/2025	14:37	LB136097
	Calcium	18500	20000	92	90 - 110	P	06/10/2025	14:37	LB136097
	Chromium	794	800	99	90 - 110	P	06/10/2025	14:37	LB136097
	Cobalt	1940	2000	97	90 - 110	P	06/10/2025	14:37	LB136097
	Copper	994	1000	99	90 - 110	P	06/10/2025	14:37	LB136097
	Iron	3770	4000	94	90 - 110	P	06/10/2025	14:37	LB136097
	Lead	3800	4000	95	90 - 110	P	06/10/2025	14:37	LB136097
	Magnesium	18700	20000	93	90 - 110	P	06/10/2025	14:37	LB136097
	Manganese	1860	2000	93	90 - 110	P	06/10/2025	14:37	LB136097
	Nickel	1930	2000	97	90 - 110	P	06/10/2025	14:37	LB136097
	Potassium	21800	20000	109	90 - 110	P	06/10/2025	14:37	LB136097
	Selenium	3880	4000	97	90 - 110	P	06/10/2025	14:37	LB136097
	Silver	978	1000	98	90 - 110	P	06/10/2025	14:37	LB136097
	Sodium	21900	20000	109	90 - 110	P	06/10/2025	14:37	LB136097
	Thallium	3740	4000	94	90 - 110	P	06/10/2025	14:37	LB136097
Vanadium	1900	2000	95	90 - 110	P	06/10/2025	14:37	LB136097	
Zinc	1980	2000	99	90 - 110	P	06/10/2025	14:37	LB136097	

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Nobis Group **SDG No.:** Q2259
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q2259 **SAS No.:** Q2259
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	110	100	110	80 - 120	P	06/10/2025	14:41	LB136097
	Antimony	55.0	50.0	110	80 - 120	P	06/10/2025	14:41	LB136097
	Arsenic	21.2	20.0	106	80 - 120	P	06/10/2025	14:41	LB136097
	Barium	101	100	101	80 - 120	P	06/10/2025	14:41	LB136097
	Beryllium	6.23	6.0	104	80 - 120	P	06/10/2025	14:41	LB136097
	Cadmium	6.31	6.0	105	80 - 120	P	06/10/2025	14:41	LB136097
	Calcium	2040	2000	102	80 - 120	P	06/10/2025	14:41	LB136097
	Chromium	10.8	10.0	108	80 - 120	P	06/10/2025	14:41	LB136097
	Cobalt	30.5	30.0	102	80 - 120	P	06/10/2025	14:41	LB136097
	Copper	22.6	20.0	113	80 - 120	P	06/10/2025	14:41	LB136097
	Iron	108	100	108	80 - 120	P	06/10/2025	14:41	LB136097
	Lead	12.8	12.0	107	80 - 120	P	06/10/2025	14:41	LB136097
	Magnesium	2030	2000	102	80 - 120	P	06/10/2025	14:41	LB136097
	Manganese	21.1	20.0	106	80 - 120	P	06/10/2025	14:41	LB136097
	Nickel	41.6	40.0	104	80 - 120	P	06/10/2025	14:41	LB136097
	Potassium	1960	2000	98	80 - 120	P	06/10/2025	14:41	LB136097
	Selenium	19.7	20.0	98	80 - 120	P	06/10/2025	14:41	LB136097
	Silver	10.5	10.0	105	80 - 120	P	06/10/2025	14:41	LB136097
	Sodium	1920	2000	96	80 - 120	P	06/10/2025	14:41	LB136097
	Thallium	41.3	40.0	103	80 - 120	P	06/10/2025	14:41	LB136097
	Vanadium	38.7	40.0	97	80 - 120	P	06/10/2025	14:41	LB136097
	Zinc	43.7	40.0	109	80 - 120	P	06/10/2025	14:41	LB136097

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Nobis Group **SDG No.:** Q2259
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q2259 **SAS No.:** Q2259
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	10100	10000	101	90 - 110	P	06/10/2025	15:30	LB136097
	Antimony	5060	5000	101	90 - 110	P	06/10/2025	15:30	LB136097
	Arsenic	5050	5000	101	90 - 110	P	06/10/2025	15:30	LB136097
	Barium	9840	10000	98	90 - 110	P	06/10/2025	15:30	LB136097
	Beryllium	262	250	105	90 - 110	P	06/10/2025	15:30	LB136097
	Cadmium	2510	2500	100	90 - 110	P	06/10/2025	15:30	LB136097
	Calcium	24900	25000	100	90 - 110	P	06/10/2025	15:30	LB136097
	Chromium	1030	1000	103	90 - 110	P	06/10/2025	15:30	LB136097
	Cobalt	2500	2500	100	90 - 110	P	06/10/2025	15:30	LB136097
	Copper	1270	1250	101	90 - 110	P	06/10/2025	15:30	LB136097
	Iron	4860	5000	97	90 - 110	P	06/10/2025	15:30	LB136097
	Lead	5030	5000	100	90 - 110	P	06/10/2025	15:30	LB136097
	Magnesium	25000	25000	100	90 - 110	P	06/10/2025	15:30	LB136097
	Manganese	2480	2500	99	90 - 110	P	06/10/2025	15:30	LB136097
	Nickel	2510	2500	100	90 - 110	P	06/10/2025	15:30	LB136097
	Potassium	24200	25000	97	90 - 110	P	06/10/2025	15:30	LB136097
	Selenium	5080	5000	102	90 - 110	P	06/10/2025	15:30	LB136097
	Silver	1270	1250	101	90 - 110	P	06/10/2025	15:30	LB136097
	Sodium	24100	25000	96	90 - 110	P	06/10/2025	15:30	LB136097
	Thallium	5100	5000	102	90 - 110	P	06/10/2025	15:30	LB136097
	Vanadium	2510	2500	100	90 - 110	P	06/10/2025	15:30	LB136097
	Zinc	2560	2500	102	90 - 110	P	06/10/2025	15:30	LB136097
CCV02	Aluminum	10000	10000	100	90 - 110	P	06/10/2025	16:28	LB136097
	Antimony	5040	5000	101	90 - 110	P	06/10/2025	16:28	LB136097
	Arsenic	5070	5000	101	90 - 110	P	06/10/2025	16:28	LB136097
	Barium	9830	10000	98	90 - 110	P	06/10/2025	16:28	LB136097
	Beryllium	261	250	104	90 - 110	P	06/10/2025	16:28	LB136097
	Cadmium	2540	2500	102	90 - 110	P	06/10/2025	16:28	LB136097
	Calcium	24900	25000	100	90 - 110	P	06/10/2025	16:28	LB136097
	Chromium	1030	1000	103	90 - 110	P	06/10/2025	16:28	LB136097
	Cobalt	2530	2500	101	90 - 110	P	06/10/2025	16:28	LB136097
	Copper	1280	1250	102	90 - 110	P	06/10/2025	16:28	LB136097
	Iron	4930	5000	99	90 - 110	P	06/10/2025	16:28	LB136097
	Lead	5070	5000	101	90 - 110	P	06/10/2025	16:28	LB136097

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Nobis Group **SDG No.:** Q2259
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q2259 **SAS No.:** Q2259
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	25000	25000	100	90 - 110	P	06/10/2025	16:28	LB136097
	Manganese	2460	2500	98	90 - 110	P	06/10/2025	16:28	LB136097
	Nickel	2540	2500	101	90 - 110	P	06/10/2025	16:28	LB136097
	Potassium	24500	25000	98	90 - 110	P	06/10/2025	16:28	LB136097
	Selenium	5100	5000	102	90 - 110	P	06/10/2025	16:28	LB136097
	Silver	1270	1250	102	90 - 110	P	06/10/2025	16:28	LB136097
	Sodium	24400	25000	97	90 - 110	P	06/10/2025	16:28	LB136097
	Thallium	5070	5000	101	90 - 110	P	06/10/2025	16:28	LB136097
	Vanadium	2510	2500	100	90 - 110	P	06/10/2025	16:28	LB136097
	Zinc	2570	2500	103	90 - 110	P	06/10/2025	16:28	LB136097
CCV03	Aluminum	9930	10000	99	90 - 110	P	06/10/2025	17:36	LB136097
	Antimony	5110	5000	102	90 - 110	P	06/10/2025	17:36	LB136097
	Arsenic	5130	5000	103	90 - 110	P	06/10/2025	17:36	LB136097
	Barium	10200	10000	102	90 - 110	P	06/10/2025	17:36	LB136097
	Beryllium	243	250	97	90 - 110	P	06/10/2025	17:36	LB136097
	Cadmium	2540	2500	102	90 - 110	P	06/10/2025	17:36	LB136097
	Calcium	25000	25000	100	90 - 110	P	06/10/2025	17:36	LB136097
	Chromium	1030	1000	103	90 - 110	P	06/10/2025	17:36	LB136097
	Cobalt	2530	2500	101	90 - 110	P	06/10/2025	17:36	LB136097
	Copper	1280	1250	102	90 - 110	P	06/10/2025	17:36	LB136097
	Iron	5390	5000	108	90 - 110	P	06/10/2025	17:36	LB136097
	Lead	5080	5000	102	90 - 110	P	06/10/2025	17:36	LB136097
	Magnesium	24700	25000	99	90 - 110	P	06/10/2025	17:36	LB136097
	Manganese	2500	2500	100	90 - 110	P	06/10/2025	17:36	LB136097
	Nickel	2540	2500	102	90 - 110	P	06/10/2025	17:36	LB136097
	Potassium	26400	25000	105	90 - 110	P	06/10/2025	17:36	LB136097
	Selenium	5140	5000	103	90 - 110	P	06/10/2025	17:36	LB136097
	Silver	1280	1250	103	90 - 110	P	06/10/2025	17:36	LB136097
	Sodium	26700	25000	107	90 - 110	P	06/10/2025	17:36	LB136097
	Thallium	4990	5000	100	90 - 110	P	06/10/2025	17:36	LB136097
	Vanadium	2540	2500	101	90 - 110	P	06/10/2025	17:36	LB136097
	Zinc	2580	2500	103	90 - 110	P	06/10/2025	17:36	LB136097
CCV04	Aluminum	9910	10000	99	90 - 110	P	06/10/2025	18:28	LB136097
	Antimony	5130	5000	103	90 - 110	P	06/10/2025	18:28	LB136097

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Nobis Group SDG No.: Q2259
 Contract: NOBI03 Lab Code: CHEM Case No.: Q2259 SAS No.: Q2259
 Initial Calibration Source: EPA
 Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	5200	5000	104	90 - 110	P	06/10/2025	18:28	LB136097
	Barium	10300	10000	103	90 - 110	P	06/10/2025	18:28	LB136097
	Beryllium	246	250	99	90 - 110	P	06/10/2025	18:28	LB136097
	Cadmium	2580	2500	103	90 - 110	P	06/10/2025	18:28	LB136097
	Calcium	25200	25000	101	90 - 110	P	06/10/2025	18:28	LB136097
	Chromium	1050	1000	105	90 - 110	P	06/10/2025	18:28	LB136097
	Cobalt	2560	2500	102	90 - 110	P	06/10/2025	18:28	LB136097
	Copper	1290	1250	103	90 - 110	P	06/10/2025	18:28	LB136097
	Iron	5420	5000	108	90 - 110	P	06/10/2025	18:28	LB136097
	Lead	5160	5000	103	90 - 110	P	06/10/2025	18:28	LB136097
	Magnesium	24800	25000	99	90 - 110	P	06/10/2025	18:28	LB136097
	Manganese	2510	2500	100	90 - 110	P	06/10/2025	18:28	LB136097
	Nickel	2570	2500	103	90 - 110	P	06/10/2025	18:28	LB136097
	Potassium	25900	25000	104	90 - 110	P	06/10/2025	18:28	LB136097
	Selenium	5220	5000	104	90 - 110	P	06/10/2025	18:28	LB136097
	Silver	1290	1250	103	90 - 110	P	06/10/2025	18:28	LB136097
	Sodium	26000	25000	104	90 - 110	P	06/10/2025	18:28	LB136097
	Thallium	5110	5000	102	90 - 110	P	06/10/2025	18:28	LB136097
	Vanadium	2550	2500	102	90 - 110	P	06/10/2025	18:28	LB136097
	Zinc	2540	2500	101	90 - 110	P	06/10/2025	18:28	LB136097
CCV05	Aluminum	9840	10000	98	90 - 110	P	06/10/2025	19:18	LB136097
	Antimony	5120	5000	102	90 - 110	P	06/10/2025	19:18	LB136097
	Arsenic	5170	5000	103	90 - 110	P	06/10/2025	19:18	LB136097
	Barium	10200	10000	102	90 - 110	P	06/10/2025	19:18	LB136097
	Beryllium	243	250	97	90 - 110	P	06/10/2025	19:18	LB136097
	Cadmium	2550	2500	102	90 - 110	P	06/10/2025	19:18	LB136097
	Calcium	24900	25000	100	90 - 110	P	06/10/2025	19:18	LB136097
	Chromium	1030	1000	103	90 - 110	P	06/10/2025	19:18	LB136097
	Cobalt	2530	2500	101	90 - 110	P	06/10/2025	19:18	LB136097
	Copper	1270	1250	102	90 - 110	P	06/10/2025	19:18	LB136097
	Iron	5340	5000	107	90 - 110	P	06/10/2025	19:18	LB136097
	Lead	5090	5000	102	90 - 110	P	06/10/2025	19:18	LB136097
	Magnesium	24400	25000	98	90 - 110	P	06/10/2025	19:18	LB136097
	Manganese	2490	2500	100	90 - 110	P	06/10/2025	19:18	LB136097

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Nobis Group **SDG No.:** Q2259
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q2259 **SAS No.:** Q2259
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2540	2500	102	90 - 110	P	06/10/2025	19:18	LB136097
	Potassium	25200	25000	101	90 - 110	P	06/10/2025	19:18	LB136097
	Selenium	5190	5000	104	90 - 110	P	06/10/2025	19:18	LB136097
	Silver	1260	1250	101	90 - 110	P	06/10/2025	19:18	LB136097
	Sodium	25200	25000	101	90 - 110	P	06/10/2025	19:18	LB136097
	Thallium	5000	5000	100	90 - 110	P	06/10/2025	19:18	LB136097
	Vanadium	2540	2500	102	90 - 110	P	06/10/2025	19:18	LB136097
	Zinc	2510	2500	100	90 - 110	P	06/10/2025	19:18	LB136097
CCV06	Aluminum	10100	10000	100	90 - 110	P	06/10/2025	20:09	LB136097
	Antimony	5120	5000	102	90 - 110	P	06/10/2025	20:09	LB136097
	Arsenic	5160	5000	103	90 - 110	P	06/10/2025	20:09	LB136097
	Barium	9960	10000	100	90 - 110	P	06/10/2025	20:09	LB136097
	Beryllium	262	250	105	90 - 110	P	06/10/2025	20:09	LB136097
	Cadmium	2550	2500	102	90 - 110	P	06/10/2025	20:09	LB136097
	Calcium	25100	25000	100	90 - 110	P	06/10/2025	20:09	LB136097
	Chromium	1050	1000	105	90 - 110	P	06/10/2025	20:09	LB136097
	Cobalt	2540	2500	102	90 - 110	P	06/10/2025	20:09	LB136097
	Copper	1270	1250	102	90 - 110	P	06/10/2025	20:09	LB136097
	Iron	5150	5000	103	90 - 110	P	06/10/2025	20:09	LB136097
	Lead	5110	5000	102	90 - 110	P	06/10/2025	20:09	LB136097
	Magnesium	25000	25000	100	90 - 110	P	06/10/2025	20:09	LB136097
	Manganese	2500	2500	100	90 - 110	P	06/10/2025	20:09	LB136097
	Nickel	2550	2500	102	90 - 110	P	06/10/2025	20:09	LB136097
	Potassium	24100	25000	96	90 - 110	P	06/10/2025	20:09	LB136097
	Selenium	5190	5000	104	90 - 110	P	06/10/2025	20:09	LB136097
	Silver	1290	1250	103	90 - 110	P	06/10/2025	20:09	LB136097
	Sodium	23800	25000	95	90 - 110	P	06/10/2025	20:09	LB136097
	Thallium	5010	5000	100	90 - 110	P	06/10/2025	20:09	LB136097
	Vanadium	2560	2500	102	90 - 110	P	06/10/2025	20:09	LB136097
	Zinc	2570	2500	103	90 - 110	P	06/10/2025	20:09	LB136097
CCV07	Aluminum	10100	10000	101	90 - 110	P	06/10/2025	21:04	LB136097
	Antimony	5160	5000	103	90 - 110	P	06/10/2025	21:04	LB136097
	Arsenic	5230	5000	104	90 - 110	P	06/10/2025	21:04	LB136097
	Barium	10100	10000	101	90 - 110	P	06/10/2025	21:04	LB136097

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Nobis Group **SDG No.:** Q2259
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q2259 **SAS No.:** Q2259
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Beryllium	261	250	104	90 - 110	P	06/10/2025	21:04	LB136097
	Cadmium	2600	2500	104	90 - 110	P	06/10/2025	21:04	LB136097
	Calcium	25600	25000	102	90 - 110	P	06/10/2025	21:04	LB136097
	Chromium	1050	1000	105	90 - 110	P	06/10/2025	21:04	LB136097
	Cobalt	2590	2500	103	90 - 110	P	06/10/2025	21:04	LB136097
	Copper	1270	1250	102	90 - 110	P	06/10/2025	21:04	LB136097
	Iron	5190	5000	104	90 - 110	P	06/10/2025	21:04	LB136097
	Lead	5210	5000	104	90 - 110	P	06/10/2025	21:04	LB136097
	Magnesium	25400	25000	102	90 - 110	P	06/10/2025	21:04	LB136097
	Manganese	2550	2500	102	90 - 110	P	06/10/2025	21:04	LB136097
	Nickel	2590	2500	104	90 - 110	P	06/10/2025	21:04	LB136097
	Potassium	23600	25000	94	90 - 110	P	06/10/2025	21:04	LB136097
	Selenium	5230	5000	105	90 - 110	P	06/10/2025	21:04	LB136097
	Silver	1280	1250	102	90 - 110	P	06/10/2025	21:04	LB136097
	Sodium	23500	25000	94	90 - 110	P	06/10/2025	21:04	LB136097
	Thallium	5090	5000	102	90 - 110	P	06/10/2025	21:04	LB136097
	Vanadium	2590	2500	104	90 - 110	P	06/10/2025	21:04	LB136097
	Zinc	2540	2500	102	90 - 110	P	06/10/2025	21:04	LB136097
CCV08	Aluminum	10100	10000	101	90 - 110	P	06/10/2025	22:04	LB136097
	Antimony	5200	5000	104	90 - 110	P	06/10/2025	22:04	LB136097
	Arsenic	5240	5000	105	90 - 110	P	06/10/2025	22:04	LB136097
	Barium	10100	10000	101	90 - 110	P	06/10/2025	22:04	LB136097
	Beryllium	262	250	105	90 - 110	P	06/10/2025	22:04	LB136097
	Cadmium	2550	2500	102	90 - 110	P	06/10/2025	22:04	LB136097
	Calcium	25400	25000	101	90 - 110	P	06/10/2025	22:04	LB136097
	Chromium	1050	1000	105	90 - 110	P	06/10/2025	22:04	LB136097
	Cobalt	2560	2500	102	90 - 110	P	06/10/2025	22:04	LB136097
	Copper	1280	1250	103	90 - 110	P	06/10/2025	22:04	LB136097
	Iron	5140	5000	103	90 - 110	P	06/10/2025	22:04	LB136097
	Lead	5130	5000	102	90 - 110	P	06/10/2025	22:04	LB136097
	Magnesium	25300	25000	101	90 - 110	P	06/10/2025	22:04	LB136097
	Manganese	2540	2500	102	90 - 110	P	06/10/2025	22:04	LB136097
	Nickel	2560	2500	103	90 - 110	P	06/10/2025	22:04	LB136097
	Potassium	23500	25000	94	90 - 110	P	06/10/2025	22:04	LB136097

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Nobis Group **SDG No.:** Q2259

Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q2259 **SAS No.:** Q2259

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV08	Selenium	5280	5000	106	90 - 110	P	06/10/2025	22:04	LB136097
	Silver	1280	1250	102	90 - 110	P	06/10/2025	22:04	LB136097
	Sodium	23400	25000	94	90 - 110	P	06/10/2025	22:04	LB136097
	Thallium	5010	5000	100	90 - 110	P	06/10/2025	22:04	LB136097
	Vanadium	2580	2500	103	90 - 110	P	06/10/2025	22:04	LB136097
	Zinc	2550	2500	102	90 - 110	P	06/10/2025	22:04	LB136097
CCV09	Aluminum	10000	10000	100	90 - 110	P	06/10/2025	22:56	LB136097
	Antimony	5110	5000	102	90 - 110	P	06/10/2025	22:56	LB136097
	Arsenic	5150	5000	103	90 - 110	P	06/10/2025	22:56	LB136097
	Barium	10100	10000	101	90 - 110	P	06/10/2025	22:56	LB136097
	Beryllium	260	250	104	90 - 110	P	06/10/2025	22:56	LB136097
	Cadmium	2550	2500	102	90 - 110	P	06/10/2025	22:56	LB136097
	Calcium	25100	25000	101	90 - 110	P	06/10/2025	22:56	LB136097
	Chromium	1050	1000	105	90 - 110	P	06/10/2025	22:56	LB136097
	Cobalt	2550	2500	102	90 - 110	P	06/10/2025	22:56	LB136097
	Copper	1270	1250	102	90 - 110	P	06/10/2025	22:56	LB136097
	Iron	5160	5000	103	90 - 110	P	06/10/2025	22:56	LB136097
	Lead	5100	5000	102	90 - 110	P	06/10/2025	22:56	LB136097
	Magnesium	25000	25000	100	90 - 110	P	06/10/2025	22:56	LB136097
	Manganese	2510	2500	100	90 - 110	P	06/10/2025	22:56	LB136097
	Nickel	2560	2500	102	90 - 110	P	06/10/2025	22:56	LB136097
	Potassium	24100	25000	96	90 - 110	P	06/10/2025	22:56	LB136097
	Selenium	5180	5000	104	90 - 110	P	06/10/2025	22:56	LB136097
	Silver	1290	1250	103	90 - 110	P	06/10/2025	22:56	LB136097
	Sodium	23900	25000	96	90 - 110	P	06/10/2025	22:56	LB136097
	Thallium	5020	5000	100	90 - 110	P	06/10/2025	22:56	LB136097
	Vanadium	2540	2500	102	90 - 110	P	06/10/2025	22:56	LB136097
	Zinc	2580	2500	103	90 - 110	P	06/10/2025	22:56	LB136097
CCV10	Aluminum	10200	10000	102	90 - 110	P	06/10/2025	23:57	LB136097
	Antimony	5170	5000	103	90 - 110	P	06/10/2025	23:57	LB136097
	Arsenic	5180	5000	104	90 - 110	P	06/10/2025	23:57	LB136097
	Barium	10000	10000	100	90 - 110	P	06/10/2025	23:57	LB136097
	Beryllium	264	250	106	90 - 110	P	06/10/2025	23:57	LB136097
	Cadmium	2540	2500	102	90 - 110	P	06/10/2025	23:57	LB136097

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Nobis Group **SDG No.:** Q2259
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q2259 **SAS No.:** Q2259
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV10	Calcium	25400	25000	102	90 - 110	P	06/10/2025	23:57	LB136097
	Chromium	1060	1000	106	90 - 110	P	06/10/2025	23:57	LB136097
	Cobalt	2550	2500	102	90 - 110	P	06/10/2025	23:57	LB136097
	Copper	1280	1250	102	90 - 110	P	06/10/2025	23:57	LB136097
	Iron	5200	5000	104	90 - 110	P	06/10/2025	23:57	LB136097
	Lead	5100	5000	102	90 - 110	P	06/10/2025	23:57	LB136097
	Magnesium	25400	25000	102	90 - 110	P	06/10/2025	23:57	LB136097
	Manganese	2530	2500	101	90 - 110	P	06/10/2025	23:57	LB136097
	Nickel	2560	2500	102	90 - 110	P	06/10/2025	23:57	LB136097
	Potassium	24000	25000	96	90 - 110	P	06/10/2025	23:57	LB136097
	Selenium	5230	5000	105	90 - 110	P	06/10/2025	23:57	LB136097
	Silver	1310	1250	104	90 - 110	P	06/10/2025	23:57	LB136097
	Sodium	23600	25000	94	90 - 110	P	06/10/2025	23:57	LB136097
	Thallium	4940	5000	99	90 - 110	P	06/10/2025	23:57	LB136097
	Vanadium	2570	2500	103	90 - 110	P	06/10/2025	23:57	LB136097
	Zinc	2630	2500	105	90 - 110	P	06/10/2025	23:57	LB136097
CCV11	Aluminum	10200	10000	102	90 - 110	P	06/11/2025	01:35	LB136097
	Antimony	5240	5000	105	90 - 110	P	06/11/2025	01:35	LB136097
	Arsenic	5280	5000	106	90 - 110	P	06/11/2025	01:35	LB136097
	Barium	10100	10000	101	90 - 110	P	06/11/2025	01:35	LB136097
	Beryllium	259	250	104	90 - 110	P	06/11/2025	01:35	LB136097
	Cadmium	2570	2500	103	90 - 110	P	06/11/2025	01:35	LB136097
	Calcium	25100	25000	100	90 - 110	P	06/11/2025	01:35	LB136097
	Chromium	1040	1000	104	90 - 110	P	06/11/2025	01:35	LB136097
	Cobalt	2580	2500	103	90 - 110	P	06/11/2025	01:35	LB136097
	Copper	1290	1250	103	90 - 110	P	06/11/2025	01:35	LB136097
	Iron	5060	5000	101	90 - 110	P	06/11/2025	01:35	LB136097
	Lead	5160	5000	103	90 - 110	P	06/11/2025	01:35	LB136097
	Magnesium	24900	25000	100	90 - 110	P	06/11/2025	01:35	LB136097
	Manganese	2520	2500	101	90 - 110	P	06/11/2025	01:35	LB136097
	Nickel	2590	2500	104	90 - 110	P	06/11/2025	01:35	LB136097
	Potassium	23800	25000	95	90 - 110	P	06/11/2025	01:35	LB136097
	Selenium	5320	5000	106	90 - 110	P	06/11/2025	01:35	LB136097
	Silver	1280	1250	103	90 - 110	P	06/11/2025	01:35	LB136097

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Nobis Group **SDG No.:** Q2259
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q2259 **SAS No.:** Q2259
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV11	Sodium	23700	25000	95	90 - 110	P	06/11/2025	01:35	LB136097
	Thallium	4950	5000	99	90 - 110	P	06/11/2025	01:35	LB136097
	Vanadium	2560	2500	102	90 - 110	P	06/11/2025	01:35	LB136097
	Zinc	2600	2500	104	90 - 110	P	06/11/2025	01:35	LB136097
CCV12	Aluminum	10200	10000	102	90 - 110	P	06/11/2025	02:21	LB136097
	Antimony	4840	5000	97	90 - 110	P	06/11/2025	02:21	LB136097
	Arsenic	4850	5000	97	90 - 110	P	06/11/2025	02:21	LB136097
	Barium	10000	10000	100	90 - 110	P	06/11/2025	02:21	LB136097
	Beryllium	264	250	105	90 - 110	P	06/11/2025	02:21	LB136097
	Cadmium	2400	2500	96	90 - 110	P	06/11/2025	02:21	LB136097
	Calcium	24900	25000	99	90 - 110	P	06/11/2025	02:21	LB136097
	Chromium	1030	1000	103	90 - 110	P	06/11/2025	02:21	LB136097
	Cobalt	2420	2500	97	90 - 110	P	06/11/2025	02:21	LB136097
	Copper	1200	1250	96	90 - 110	P	06/11/2025	02:21	LB136097
	Iron	4930	5000	99	90 - 110	P	06/11/2025	02:21	LB136097
	Lead	4810	5000	96	90 - 110	P	06/11/2025	02:21	LB136097
	Magnesium	24900	25000	100	90 - 110	P	06/11/2025	02:21	LB136097
	Manganese	2490	2500	100	90 - 110	P	06/11/2025	02:21	LB136097
	Nickel	2420	2500	97	90 - 110	P	06/11/2025	02:21	LB136097
	Potassium	23100	25000	92	90 - 110	P	06/11/2025	02:21	LB136097
	Selenium	4880	5000	98	90 - 110	P	06/11/2025	02:21	LB136097
	Silver	1280	1250	102	90 - 110	P	06/11/2025	02:21	LB136097
	Sodium	22800	25000	91	90 - 110	P	06/11/2025	02:21	LB136097
	Thallium	4660	5000	93	90 - 110	P	06/11/2025	02:21	LB136097
	Vanadium	2530	2500	101	90 - 110	P	06/11/2025	02:21	LB136097
	Zinc	2590	2500	104	90 - 110	P	06/11/2025	02:21	LB136097