



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Nobis Group SDG No.: Q2259
Contract: NOBI03 Lab Code: CHEM Case No.: Q2259 SAS No.: Q2259
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.19	0.2	93	70 - 130	CV	06/12/2025	10:21	LB136124
CRI	Antimony	1.81	2.0	90	70 - 130	P	06/20/2025	12:14	LB136217
	Arsenic	1.03	1.0	103	70 - 130	P	06/20/2025	12:14	LB136217
	Barium	8.57	10.0	86	70 - 130	P	06/20/2025	12:14	LB136217
	Beryllium	0.99	1.0	99	70 - 130	P	06/20/2025	12:14	LB136217
	Cadmium	0.91	1.0	91	70 - 130	P	06/20/2025	12:14	LB136217
	Chromium	1.87	2.0	94	70 - 130	P	06/20/2025	12:14	LB136217
	Copper	1.93	2.0	96	70 - 130	P	06/20/2025	12:14	LB136217
	Lead	0.77	1.0	77	70 - 130	P	06/20/2025	12:14	LB136217
	Nickel	0.99	1.0	99	70 - 130	P	06/20/2025	12:14	LB136217
	Selenium	4.31	5.0	86	70 - 130	P	06/20/2025	12:14	LB136217
	Silver	1.00	1.0	100	70 - 130	P	06/20/2025	12:14	LB136217
	Thallium	0.90	1.0	90	70 - 130	P	06/20/2025	12:14	LB136217
	Vanadium	4.38	5.0	88	70 - 130	P	06/20/2025	12:14	LB136217
	Zinc	5.32	5.0	106	50 - 150	P	06/20/2025	12:14	LB136217