



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Nobis Group				SDG No.:	Q2259				
Contract:	NOBI03	Lab Code:	CHEM	Case No.:	Q2259	SAS No.:	Q2259			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB35	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	06/12/2025	10:14	LB136124

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group **SDG No.:** Q2259
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q2259 **SAS No.:** Q2259

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB30	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	06/12/2025	10:19	LB136124
CCB31	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	06/12/2025	10:52	LB136124
CCB32	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	06/12/2025	11:20	LB136124
CCB33	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	06/12/2025	11:40	LB136124
CCB34	Mercury	0.076	+/-0.2	U	0.16	0.20	CV	06/12/2025	12:01	LB136124

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group SDG No.: Q2259
Contract: NOBI03 Lab Code: CHEM Case No.: Q2259 SAS No.: Q2259

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Antimony	0.11	+/-1	U	0.25	2.00	P	06/20/2025	11:42	LB136217
	Arsenic	0.089	+/-0.5	U	0.25	1.00	P	06/20/2025	11:42	LB136217
	Barium	0.21	+/-5	U	1.25	10.0	P	06/20/2025	11:42	LB136217
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	06/20/2025	11:42	LB136217
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	06/20/2025	11:42	LB136217
	Chromium	0.21	+/-1	U	0.75	2.00	P	06/20/2025	11:42	LB136217
	Copper	0.30	+/-1	U	1.50	2.00	P	06/20/2025	11:42	LB136217
	Lead	0.21	+/-0.5	U	0.75	1.00	P	06/20/2025	11:42	LB136217
	Nickel	0.27	+/-0.5	U	0.75	1.00	P	06/20/2025	11:42	LB136217
	Selenium	2.90	+/-2.5	U	4.50	5.00	P	06/20/2025	11:42	LB136217
	Silver	0.060	+/-0.5	U	0.50	1.00	P	06/20/2025	11:42	LB136217
	Thallium	0.060	+/-0.5	U	0.50	1.00	P	06/20/2025	11:42	LB136217
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	06/20/2025	11:42	LB136217
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	06/20/2025	11:42	LB136217

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Nobis Group		SDG No.: Q2259								
Contract:	NOBI03	Lab Code:	CHEM	Case No.: Q2259			SAS No.: Q2259			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Antimony	0.11	+/-1	U	0.25	2.00	P	06/20/2025	12:11	LB136217
	Arsenic	0.089	+/-0.5	U	0.25	1.00	P	06/20/2025	12:11	LB136217
	Barium	0.21	+/-5	U	1.25	10.0	P	06/20/2025	12:11	LB136217
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	06/20/2025	12:11	LB136217
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	06/20/2025	12:11	LB136217
	Chromium	0.21	+/-1	U	0.75	2.00	P	06/20/2025	12:11	LB136217
	Copper	0.30	+/-1	U	1.50	2.00	P	06/20/2025	12:11	LB136217
	Lead	0.21	+/-0.5	U	0.75	1.00	P	06/20/2025	12:11	LB136217
	Nickel	0.27	+/-0.5	U	0.75	1.00	P	06/20/2025	12:11	LB136217
	Selenium	2.90	+/-2.5	U	4.50	5.00	P	06/20/2025	12:11	LB136217
	Silver	0.060	+/-0.5	U	0.50	1.00	P	06/20/2025	12:11	LB136217
	Thallium	0.060	+/-0.5	J	0.50	1.00	P	06/20/2025	12:11	LB136217
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	06/20/2025	12:11	LB136217
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	06/20/2025	12:11	LB136217
CCB02	Antimony	0.11	+/-1	U	0.25	2.00	P	06/20/2025	12:49	LB136217
	Arsenic	0.089	+/-0.5	U	0.25	1.00	P	06/20/2025	12:49	LB136217
	Barium	0.21	+/-5	U	1.25	10.0	P	06/20/2025	12:49	LB136217
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	06/20/2025	12:49	LB136217
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	06/20/2025	12:49	LB136217
	Chromium	0.21	+/-1	U	0.75	2.00	P	06/20/2025	12:49	LB136217
	Copper	0.30	+/-1	U	1.50	2.00	P	06/20/2025	12:49	LB136217
	Lead	0.21	+/-0.5	U	0.75	1.00	P	06/20/2025	12:49	LB136217
	Nickel	0.27	+/-0.5	U	0.75	1.00	P	06/20/2025	12:49	LB136217
	Selenium	2.90	+/-2.5	U	4.50	5.00	P	06/20/2025	12:49	LB136217
	Silver	0.060	+/-0.5	U	0.50	1.00	P	06/20/2025	12:49	LB136217
	Thallium	0.060	+/-0.5	U	0.50	1.00	P	06/20/2025	12:49	LB136217
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	06/20/2025	12:49	LB136217
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	06/20/2025	12:49	LB136217
CCB03	Antimony	0.11	+/-1	J	0.25	2.00	P	06/20/2025	13:25	LB136217
	Arsenic	0.089	+/-0.5	U	0.25	1.00	P	06/20/2025	13:25	LB136217
	Barium	0.21	+/-5	U	1.25	10.0	P	06/20/2025	13:25	LB136217
	Beryllium	0.32	+/-0.5	U	0.75	1.00	P	06/20/2025	13:25	LB136217
	Cadmium	0.34	+/-0.5	U	0.50	1.00	P	06/20/2025	13:25	LB136217
	Chromium	0.21	+/-1	U	0.75	2.00	P	06/20/2025	13:25	LB136217
	Copper	0.30	+/-1	U	1.50	2.00	P	06/20/2025	13:25	LB136217
	Lead	0.21	+/-0.5	U	0.75	1.00	P	06/20/2025	13:25	LB136217
	Nickel	0.27	+/-0.5	U	0.75	1.00	P	06/20/2025	13:25	LB136217
	Selenium	2.90	+/-2.5	U	4.50	5.00	P	06/20/2025	13:25	LB136217
	Silver	0.060	+/-0.5	U	0.50	1.00	P	06/20/2025	13:25	LB136217

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>Nobis Group</u>				SDG No.:	<u>Q2259</u>				
Contract:	<u>NOBI03</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q2259</u>	SAS No.:	<u>Q2259</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB03	Thallium	0.060	+/-0.5	U	0.50	1.00	P	06/20/2025	13:25	LB136217
	Vanadium	0.077	+/-2.5	U	0.25	5.00	P	06/20/2025	13:25	LB136217
	Zinc	1.25	+/-2.5	U	1.50	5.00	P	06/20/2025	13:25	LB136217