

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2259

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168424BL		WATER		Batch Number:		PB168424		Prep Date:	06/11/2025	
	Mercury	0.076	<0.2	U	0.16	0.20	CV	06/12/2025	10:28	LB136124
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168424TB		WATER		Batch Number:		PB168424		Prep Date:	06/11/2025	
	Mercury	0.076	<0.2	U	0.16	0.20	CV	06/12/2025	11:42	LB136124

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Nobis Group

SDG No.: Q2259

Instrument: P7

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168389BL	WATER			Batch Number:		PB168389		Prep Date:	06/10/2025	
	Antimony	0.11	<1	U	0.25	2.00	P	06/20/2025	12:17	LB136217
	Arsenic	0.089	<0.5	U	0.25	1.00	P	06/20/2025	12:17	LB136217
	Barium	0.21	<5	U	1.25	10.0	P	06/20/2025	12:17	LB136217
	Beryllium	0.32	<0.5	U	0.75	1.00	P	06/20/2025	12:17	LB136217
	Cadmium	0.34	<0.5	U	0.50	1.00	P	06/20/2025	12:17	LB136217
	Chromium	0.21	<1	U	0.75	2.00	P	06/20/2025	12:17	LB136217
	Copper	0.30	<1	U	1.50	2.00	P	06/20/2025	12:17	LB136217
	Lead	0.21	<0.5	U	0.75	1.00	P	06/20/2025	12:17	LB136217
	Nickel	0.27	<0.5	U	0.75	1.00	P	06/20/2025	12:17	LB136217
	Selenium	2.90	<2.5	U	4.50	5.00	P	06/20/2025	12:17	LB136217
	Silver	0.060	<0.5	U	0.50	1.00	P	06/20/2025	12:17	LB136217
	Thallium	0.060	<0.5	U	0.50	1.00	P	06/20/2025	12:17	LB136217
	Vanadium	0.077	<2.5	U	0.25	5.00	P	06/20/2025	12:17	LB136217
	Zinc	1.25	<2.5	U	1.50	5.00	P	06/20/2025	12:17	LB136217
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	LOD ug/L	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB168389TB	WATER			Batch Number:		PB168389		Prep Date:	06/10/2025	
	Antimony	0.11	<1	U	0.25	2.00	P	06/20/2025	12:23	LB136217
	Arsenic	0.089	<0.5	U	0.25	1.00	P	06/20/2025	12:23	LB136217
	Barium	0.21	<5	U	1.25	10.0	P	06/20/2025	12:23	LB136217
	Beryllium	0.32	<0.5	U	0.75	1.00	P	06/20/2025	12:23	LB136217
	Cadmium	0.34	<0.5	U	0.50	1.00	P	06/20/2025	12:23	LB136217
	Chromium	0.21	<1	U	0.75	2.00	P	06/20/2025	12:23	LB136217
	Copper	0.30	<1	U	1.50	2.00	P	06/20/2025	12:23	LB136217
	Lead	0.21	<0.5	U	0.75	1.00	P	06/20/2025	12:23	LB136217
	Nickel	0.27	<0.5	U	0.75	1.00	P	06/20/2025	12:23	LB136217
	Selenium	2.90	<2.5	U	4.50	5.00	P	06/20/2025	12:23	LB136217
	Silver	0.060	<0.5	U	0.50	1.00	P	06/20/2025	12:23	LB136217
	Thallium	0.060	<0.5	U	0.50	1.00	P	06/20/2025	12:23	LB136217
	Vanadium	0.077	<2.5	U	0.25	5.00	P	06/20/2025	12:23	LB136217
	Zinc	1.25	<2.5	U	1.50	5.00	P	06/20/2025	12:23	LB136217