

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Nobis Group **Level:** LOW **SDG No.:** Q2259
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q2259 **SAS No.:** Q2259
Matrix: Water **Sample ID:** Q2259-04 **Client ID:** OU4-PCS-TC-37-060525DUP
Percent Solids for Sample: NA **Duplicate ID** Q2259-04DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Antimony	ug/L	20	10.0	UD	10.0	UD			P
Arsenic	ug/L	20	2.10	JD	1.85	JD	13		P
Barium	ug/L	20	94.6	D	95.5	D	1		P
Beryllium	ug/L	20	5.00	UD	5.00	UD			P
Cadmium	ug/L	20	5.00	UD	5.00	UD			P
Chromium	ug/L	20	7.10	JD	6.10	JD	15		P
Copper	ug/L	20	95.0	D	52.7	D	57	*	P
Lead	ug/L	20	5.00	UD	5.00	UD			P
Nickel	ug/L	20	61.7	D	60.9	D	1		P
Mercury	ug/L	20	0.20	U	0.20	U			CV
Selenium	ug/L	20	25.0	UD	25.0	UD			P
Silver	ug/L	20	5.00	UD	5.00	UD			P
Thallium	ug/L	20	5.00	UD	5.00	UD			P
Vanadium	ug/L	20	259	D	259	D	0		P
Zinc	ug/L	20	726	D	727	D	0		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Nobis Group **Level:** LOW **SDG No.:** Q2259
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q2259 **SAS No.:** Q2259
Matrix: Water **Sample ID:** Q2259-04MS **Client ID:** OU4-PCS-TC-37-060525MSD
Percent Solids for Sample: NA **Duplicate ID** Q2259-04MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Antimony	ug/L	20	450	D	448	D	0		P
Arsenic	ug/L	20	526	D	473	D	11		P
Barium	ug/L	20	2030	D	2020	D	0		P
Beryllium	ug/L	20	457	D	453	D	1		P
Cadmium	ug/L	20	490	D	486	D	1		P
Chromium	ug/L	20	452	D	447	D	1		P
Copper	ug/L	20	4990	D	4940	D	1		P
Lead	ug/L	20	2150	D	2150	D	0		P
Nickel	ug/L	20	544	D	548	D	1		P
Mercury	ug/L	20	3.41		3.20		6		CV
Selenium	ug/L	20	441	D	446	D	1		P
Silver	ug/L	20	82.0	D	81.9	D	0		P
Thallium	ug/L	20	404	D	406	D	0		P
Vanadium	ug/L	20	705	D	699	D	1		P
Zinc	ug/L	20	5230	D	5240	D	0		P