

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Nobis Group **SDG No.:** Q2259
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q2259 **SAS No.:** Q2259

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168389BS							
Antimony	ug/L	500	498		100	85 - 117	P
Arsenic	ug/L	500	495		99	84 - 116	P
Barium	ug/L	2500	2490		100	86 - 114	P
Beryllium	ug/L	500	491		98	83 - 121	P
Cadmium	ug/L	500	505		101	87 - 115	P
Chromium	ug/L	500	466		93	85 - 116	P
Copper	ug/L	5000	5010		100	85 - 118	P
Lead	ug/L	2500	2420		97	88 - 115	P
Nickel	ug/L	500	504		101	85 - 117	P
Selenium	ug/L	500	498		100	80 - 120	P
Silver	ug/L	500	509		102	85 - 116	P
Thallium	ug/L	500	488		98	82 - 116	P
Vanadium	ug/L	500	461		92	86 - 115	P
Zinc	ug/L	5000	5120		102	83 - 119	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Nobis Group **SDG No.:** Q2259
Contract: NOBI03 **Lab Code:** CHEM **Case No.:** Q2259 **SAS No.:** Q2259

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168424BS Mercury	ug/L	4.0	3.55		89	82 - 119	CV