

FORM 8A

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: CHEM Case no.: Q2259

Sas No.: Q2259 SDG No.: Q2259

Instrument ID: P7

Start Date : 06/20/2025

Run Number: LB136217

End Date : 06/20/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 6Li	Q	Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q
S0	S0	1030	100		100		100		100		100	
S2	S2	1034	100		100		101		100		99	
S3	S3	1040	102		100		100		100		99	
S4	S4	1044	100		96		97		97		97	
S5	S5	1047	100		94		95		96		99	
S6	S6	1050	103		96		99		97		102	
S7	S7	1053	103		98		101		98		104	
S8	S8	1055	100		100		101		92		100	
ICV01	ICV01	1131	98		105		105		104		102	
LLICV01	LLICV01	1139	97		102		103		101		98	
ICB01	ICB01	1142	97		100		101		100		98	
ICSA01	ICSA01	1146	102		102		103		99		103	
ICSAB01	ICSAB01	1203	97		100		102		99		103	
CCV01	CCV01	1208	95		99		99		92		98	
CCB01	CCB01	1211	96		98		100		100		100	
CRI	CRI	1214	95		99		101		100		100	
PB168389BL	PB168389BL	1217	97		102		102		101		101	
PB168389BS	PB168389BS	1221	98		102		103		101		103	
PB168389TB	PB168389TB	1223	97		99		101		101		100	
Q2259-02	OU4-PCS-TC-	1227	98		103		113		102		104	
Q2259-04	OU4-PCS-TC-	1230	98		103		114		103		105	
Q2259-04DUP	OU4-PCS-TC-	1233	98		102		112		100		104	
Q2259-04L	OU4-PCS-TC-	1236	97		101		105		100		101	
CCV02	CCV02	1246	98		99		99		94		99	
CCB02	CCB02	1249	101		100		102		101		102	
Q2259-04MS	OU4-PCS-TC-	1302	99		101		112		101		103	
Q2259-04MSD	OU4-PCS-TC-	1305	96		98		108		98		100	
Q2259-04A	OU4-PCS-TC-	1319	98		102		114		102		104	
CCV03	CCV03	1323	98		102		103		96		103	
CCB03	CCB03	1325	99		100		102		100		100	

FORM 8A

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Nobis Group

Contract: NOBI03

Lab Code: CHEM Case no.: Q2259

Sas No.: Q2259 SDG No.: Q2259

Instrument ID: P7

Start Date : 06/20/2025

Run Number: LB136217

End Date : 06/20/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 45Sc	Q	Element 89Y	Q	Element 103Rh	Q	Element 159Tb	Q	Element 165Ho	Q
S0	S0	1030	100		100		100		100		100	
S2	S2	1034	101		100		101		101		100	
S3	S3	1040	102		101		100		100		100	
S4	S4	1044	99		98		99		99		100	
S5	S5	1047	98		96		96		99		98	
S6	S6	1050	102		100		99		102		102	
S7	S7	1053	102		102		100		104		103	
S8	S8	1055	104		103		94		101		103	
ICV01	ICV01	1131	104		107		105		104		104	
LLICV01	LLICV01	1139	101		102		103		100		100	
ICB01	ICB01	1142	101		103		103		99		100	
ICSA01	ICSA01	1146	103		103		101		103		103	
ICSAB01	ICSAB01	1203	101		103		101		103		102	
CCV01	CCV01	1208	100		100		97		100		100	
CCB01	CCB01	1211	100		102		103		101		100	
CRI	CRI	1214	102		105		104		102		102	
PB168389BL	PB168389BL	1217	101		103		105		101		102	
PB168389BS	PB168389BS	1221	103		104		102		102		102	
PB168389TB	PB168389TB	1223	99		101		103		100		100	
Q2259-02	OU4-PCS-TC-	1227	103		113		104		104		104	
Q2259-04	OU4-PCS-TC-	1230	103		113		103		104		103	
Q2259-04DUP	OU4-PCS-TC-	1233	102		112		104		103		103	
Q2259-04L	OU4-PCS-TC-	1236	101		105		103		102		102	
CCV02	CCV02	1246	101		102		96		100		99	
CCB02	CCB02	1249	101		104		104		102		101	
Q2259-04MS	OU4-PCS-TC-	1302	100		110		102		103		101	
Q2259-04MSD	OU4-PCS-TC-	1305	99		109		100		101		100	
Q2259-04A	OU4-PCS-TC-	1319	101		112		103		103		103	
CCV03	CCV03	1323	103		102		98		102		102	
CCB03	CCB03	1325	101		104		103		103		102	