

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

OU4-PCS-TC-37-060525L

Lab Name: Chemtech Consulting Group

Contract: NOBI03

Lab Code: CHEM **Lb No.:** lb136217

Lab Sample ID : Q2259-04L

SDG No.: Q2259

Matrix (soil/water): Water

Level (low/med): LOW

Concentration Units: ug/L

Analyte	Initial Sample Result (I)		Serial Dilution Result (S)		% Difference	Q	M
		C		C			
Antimony	10.0	UD	50.0	UD			P
Arsenic	2.10	JD	2.25	JD	7		P
Barium	94.6	D	98.5	JD	4		P
Beryllium	5.00	UD	25.0	UD			P
Cadmium	5.00	UD	25.0	UD			P
Chromium	7.10	JD	6.75	JD	5		P
Copper	95.0	D	77.0	D	19		P
Lead	5.00	UD	25.0	UD			P
Nickel	61.7	D	64.5	D	5		P
Mercury	0.20	U	1.00	U			CV
Selenium	25.0	UD	125	UD			P
Silver	5.00	UD	25.0	UD			P
Thallium	5.00	UD	25.0	UD			P
Vanadium	259	D	270	D	4		P
Zinc	726	D	770	D	6		P