

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q2259 SAS No.: Q2259 SDG NO.: Q2259

Continuing Calib Date: 06/17/2025 Initial Calibration Date(s): 06/16/2025 06/16/2025

Continuing Calib Time: 01:26 Initial Calibration Time(s): 09:56 11:33

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.54	7.54	7.44	7.64	0.00
2,4-DCAA	7.35	7.35	7.25	7.45	0.00
Dalapon	2.71	2.71	2.61	2.81	0.00
DICHLORPROP	8.26	8.26	8.16	8.36	0.00
2,4-D	8.49	8.49	8.39	8.59	0.00
2,4,5-TP(Silvex)	9.38	9.38	9.28	9.48	0.00
2,4,5-T	9.67	9.67	9.57	9.77	0.00
2,4-DB	10.25	10.25	10.15	10.35	0.00
Dinoseb	11.47	11.47	11.37	11.57	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q2259 SAS No.: Q2259 SDG NO.: Q2259

Continuing Calib Date: 06/17/2025 Initial Calibration Date(s): 06/16/2025 06/16/2025

Continuing Calib Time: 01:26 Initial Calibration Time(s): 09:56 11:33

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.97	7.97	7.87	8.07	0.00
2,4-DCAA	7.77	7.77	7.67	7.87	0.00
Dalapon	2.71	2.71	2.61	2.81	0.00
DICHLORPROP	8.69	8.69	8.59	8.79	0.00
2,4-D	9.03	9.03	8.93	9.13	0.01
2,4,5-TP(Silvex)	9.93	9.93	9.83	10.03	0.00
2,4,5-T	10.36	10.36	10.26	10.46	0.00
2,4-DB	10.93	10.93	10.83	11.03	0.00
Dinoseb	11.31	11.31	11.21	11.41	0.00