

Surrogate Summary

SDG No.: **Q2259**

Client: **Nobis Group**

Analytical Method: **8151A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PS030475.D	PIBLK-PS030475.D	2,4-DCAA	1	500	414	83		32	138
		2,4-DCAA	2	500	445	89		32	138
I.BLK-PS030593.D	PIBLK-PS030593.D	2,4-DCAA	1	500	459	92		32	138
		2,4-DCAA	2	500	475	95		32	138
Q2259-01	OU4-PCS-TC-36-060525	2,4-DCAA	1	500	303	61		27	122
		2,4-DCAA	2	500	321	64		27	122
Q2259-03	OU4-PCS-TC-37-060525	2,4-DCAA	1	500	154	31		27	122
		2,4-DCAA	2	500	171	34		27	122
I.BLK-PS030605.D	PIBLK-PS030605.D	2,4-DCAA	1	500	452	90		32	138
		2,4-DCAA	2	500	481	96		32	138
I.BLK-PS030620.D	PIBLK-PS030620.D	2,4-DCAA	1	500	449	90		32	138
		2,4-DCAA	2	500	487	97		32	138
Q2266-01MS	WC-3MS	2,4-DCAA	1	500	542	108		27	122
		2,4-DCAA	2	500	604	121		27	122
Q2266-01MSD	WC-3MSD	2,4-DCAA	1	500	544	109		27	122
		2,4-DCAA	2	500	610	122		27	122
PB168375BS	PB168375BS	2,4-DCAA	1	500	499	100		27	122
		2,4-DCAA	2	500	486	97		27	122
I.BLK-PS030629.D	PIBLK-PS030629.D	2,4-DCAA	1	500	452	90		32	138
		2,4-DCAA	2	500	492	98		32	138
I.BLK-PS030680.D	PIBLK-PS030680.D	2,4-DCAA	1	500	396	79		32	138
		2,4-DCAA	2	500	472	94		32	138
I.BLK-PS030687.D	PIBLK-PS030687.D	2,4-DCAA	1	500	401	80		32	138
		2,4-DCAA	2	500	473	95		32	138
PB168375BL	PB168375BL	2,4-DCAA	1	500	384	77		27	122
		2,4-DCAA	2	500	404	81		27	122
I.BLK-PS030697.D	PIBLK-PS030697.D	2,4-DCAA	1	500	446	89		32	138
		2,4-DCAA	2	500	477	95		32	138