

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: NOBI03
Lab Code: CHEM **Case No.:** Q2259 **SAS No.:** Q2259 **SDG NO.:** Q2259
Instrument ID: ECD_S **Calibration Date(s):** 06/04/2025 06/04/2025
Calibration Times: 11:19 12:55

GC Column: RTX-CLP **ID:** 0.32 (mm)

LAB FILE ID:CF 200 = PS030476.DCF 500 = PS030477.D CF 750 = PS030478.DCF 1000 = PS030479.DCF 1500 = PS030480.D							
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-T	20110200000	18445100000	18262000000	18165800000	17172200000	18431100000	6
2,4,5-TP(Silvex)	23444500000	20930200000	20252400000	19918900000	18779400000	20665100000	8
2,4-D	4208920000	3650560000	3531620000	3494410000	3332550000	3643610000	9
2,4-DB	3187030000	2969900000	2987600000	3055150000	2982130000	3036360000	3
2,4-DCAA	4505990000	3745160000	3587820000	3627070000	3409690000	3775150000	11
Dalapon	6635610000	5627170000	5409030000	5378230000	5184820000	5646970000	10
DICAMBA	17545300000	15532100000	14992100000	14866500000	14176500000	15422500000	8
DICHLORPROP	4363010000	3637540000	3480040000	3430450000	3279680000	3638150000	12
Dinoseb	16280300000	14486300000	14010300000	13907800000	13190000000	14374900000	8

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: NOBI03

Lab Code: CHEM Case No.: Q2259 SAS No.: Q2259 SDG NO.: Q2259

Instrument ID: ECD_S Calibration Date(s): 06/04/2025 06/04/2025
Calibration Times: 11:19 12:55

GC Column: RTX-CLP2 ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS030476.D</u>	CF 500 =	<u>PS030477.D</u>		
	CF 750 =	<u>PS030478.D</u>	CF 1000 =	<u>PS030479.D</u>	CF 1500 =	<u>PS030480.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-T	14777200000	13128600000	12724600000	12611900000	12043300000	13057100000	8
2,4,5-TP(Silvex)	15627400000	13837500000	13337200000	13203400000	12578300000	13716700000	8
2,4-D	1858730000	1575260000	1504430000	1488570000	1434610000	1572320000	11
2,4-DB	1329650000	1149370000	1105700000	1098150000	1064880000	1149550000	9
2,4-DCAA	1341960000	1066160000	1011260000	997042000	961634000	1075610000	14
Dalapon	3198400000	2714060000	2628110000	2611790000	2547720000	2740010000	10
DICAMBA	7190560000	6384370000	6250780000	6263160000	6119370000	6441650000	7
DICHLORPROP	1789880000	1480180000	1413150000	1395780000	1342020000	1484200000	12
Dinoseb	11251500000	10033800000	9761280000	9784360000	9419650000	10050100000	7