

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q2259 SAS No.: Q2259 SDG NO.: Q2259

Continuing Calib Date: 06/10/2025 Initial Calibration Date(s): 06/04/2025 06/04/2025

Continuing Calib Time: 20:01 Initial Calibration Time(s): 11:19 12:55

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.55	7.56	7.46	7.66	0.01
2,4-DCAA	7.36	7.37	7.27	7.47	0.01
Dalapon	2.72	2.72	2.62	2.82	0.00
DICHLORPROP	8.27	8.28	8.18	8.38	0.01
2,4-D	8.50	8.51	8.41	8.61	0.01
2,4,5-TP(Silvex)	9.39	9.40	9.30	9.50	0.01
2,4,5-T	9.68	9.70	9.60	9.80	0.02
2,4-DB	10.26	10.28	10.18	10.38	0.02
Dinoseb	11.48	11.50	11.40	11.60	0.02

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q2259 SAS No.: Q2259 SDG NO.: Q2259

Continuing Calib Date: 06/10/2025 Initial Calibration Date(s): 06/04/2025 06/04/2025

Continuing Calib Time: 20:01 Initial Calibration Time(s): 11:19 12:55

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.97	7.98	7.88	8.08	0.01
2,4-DCAA	7.77	7.77	7.67	7.87	0.00
Dalapon	2.71	2.71	2.61	2.81	0.00
DICHLORPROP	8.69	8.70	8.60	8.80	0.01
2,4-D	9.03	9.03	8.93	9.13	0.00
2,4,5-TP(Silvex)	9.93	9.94	9.84	10.04	0.01
2,4,5-T	10.36	10.37	10.27	10.47	0.01
2,4-DB	10.93	10.94	10.84	11.04	0.01
Dinoseb	11.31	11.32	11.22	11.42	0.01