

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q2259 SAS No.: Q2259 SDG NO.: Q2259

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL01 Date Analyzed: 06/10/2025

Lab Sample No.: HSTDCCC750 Data File : PS030594.D Time Analyzed: 12:27

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.683	9.597	9.797	707.270	712.500	-0.7
2,4,5-TP(Silvex)	9.389	9.304	9.504	739.160	712.500	3.7
2,4-D	8.499	8.412	8.612	693.550	705.000	-1.6
2,4-DB	10.263	10.179	10.379	652.380	712.500	-8.4
2,4-DCAA	7.360	7.269	7.469	752.900	750.000	0.4
Dalapon	2.712	2.616	2.816	709.680	682.500	4.0
DICAMBA	7.551	7.461	7.661	741.410	705.000	5.2
DICHLORPROP	8.266	8.179	8.379	713.240	705.000	1.2
Dinoseb	11.485	11.402	11.602	736.750	705.000	4.5

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q2259 SAS No.: Q2259 SDG NO.: Q2259

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL01 Date Analyzed: 06/10/2025

Lab Sample No.: HSTDCCC750 Data File : PS030594.D Time Analyzed: 12:27

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.360	10.268	10.468	729.720	712.500	2.4
2,4,5-TP(Silvex)	9.935	9.842	10.042	735.940	712.500	3.3
2,4-D	9.027	8.934	9.134	718.450	705.000	1.9
2,4-DB	10.928	10.836	11.036	707.420	712.500	-0.7
2,4-DCAA	7.769	7.674	7.874	746.710	750.000	-0.4
Dalapon	2.712	2.610	2.810	681.040	682.500	-0.2
DICAMBA	7.972	7.877	8.077	730.310	705.000	3.6
DICHLORPROP	8.692	8.598	8.798	710.110	705.000	0.7
Dinoseb	11.313	11.221	11.421	741.330	705.000	5.2