

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q2259 SAS No.: Q2259 SDG NO.: Q2259

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL02 Date Analyzed: 06/10/2025

Lab Sample No.: HSTDCCC750 Data File : PS030606.D Time Analyzed: 20:01

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.683	9.597	9.797	708.690	712.500	-0.5
2,4,5-TP(Silvex)	9.388	9.304	9.504	734.910	712.500	3.1
2,4-D	8.498	8.412	8.612	699.570	705.000	-0.8
2,4-DB	10.262	10.179	10.379	662.060	712.500	-7.1
2,4-DCAA	7.358	7.269	7.469	695.690	750.000	-7.2
Dalapon	2.716	2.616	2.816	699.180	682.500	2.4
DICAMBA	7.550	7.461	7.661	736.770	705.000	4.5
DICHLORPROP	8.265	8.179	8.379	709.770	705.000	0.7
Dinoseb	11.483	11.402	11.602	730.610	705.000	3.6

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q2259 SAS No.: Q2259 SDG NO.: Q2259

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL02 Date Analyzed: 06/10/2025

Lab Sample No.: HSTDCCC750 Data File : PS030606.D Time Analyzed: 20:01

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.360	10.268	10.468	734.330	712.500	3.1
2,4,5-TP(Silvex)	9.934	9.842	10.042	739.790	712.500	3.8
2,4-D	9.027	8.934	9.134	724.670	705.000	2.8
2,4-DB	10.927	10.836	11.036	717.750	712.500	0.7
2,4-DCAA	7.769	7.674	7.874	736.310	750.000	-1.8
Dalapon	2.711	2.610	2.810	679.800	682.500	-0.4
DICAMBA	7.972	7.877	8.077	732.020	705.000	3.8
DICHLORPROP	8.692	8.598	8.798	712.380	705.000	1.0
Dinoseb	11.313	11.221	11.421	743.840	705.000	5.5