

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q2259 SAS No.: Q2259 SDG NO.: Q2259

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL03 Date Analyzed: 06/11/2025

Lab Sample No.: HSTDCCC750 Data File : PS030621.D Time Analyzed: 09:10

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.680	9.597	9.797	700.660	712.500	-1.7
2,4,5-TP(Silvex)	9.386	9.304	9.504	736.560	712.500	3.4
2,4-D	8.496	8.412	8.612	691.670	705.000	-1.9
2,4-DB	10.259	10.179	10.379	643.460	712.500	-9.7
2,4-DCAA	7.356	7.269	7.469	737.450	750.000	-1.7
Dalapon	2.715	2.616	2.816	690.220	682.500	1.1
DICAMBA	7.549	7.461	7.661	744.120	705.000	5.5
DICHLORPROP	8.264	8.179	8.379	713.190	705.000	1.2
Dinoseb	11.481	11.402	11.602	718.070	705.000	1.9

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q2259 SAS No.: Q2259 SDG NO.: Q2259

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL03 Date Analyzed: 06/11/2025

Lab Sample No.: HSTDCCC750 Data File : PS030621.D Time Analyzed: 09:10

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.359	10.268	10.468	750.020	712.500	5.3
2,4,5-TP(Silvex)	9.932	9.842	10.042	754.780	712.500	5.9
2,4-D	9.026	8.934	9.134	740.500	705.000	5.0
2,4-DB	10.925	10.836	11.036	732.100	712.500	2.8
2,4-DCAA	7.768	7.674	7.874	751.350	750.000	0.2
Dalapon	2.711	2.610	2.810	681.600	682.500	-0.1
DICAMBA	7.971	7.877	8.077	750.620	705.000	6.5
DICHLORPROP	8.691	8.598	8.798	728.610	705.000	3.3
Dinoseb	11.311	11.221	11.421	752.930	705.000	6.8