

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q2259 SAS No.: Q2259 SDG NO.: Q2259

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL04 Date Analyzed: 06/11/2025

Lab Sample No.: HSTDCCC750 Data File : PS030630.D Time Analyzed: 15:32

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.681	9.597	9.797	676.850	712.500	-5.0
2,4,5-TP(Silvex)	9.387	9.304	9.504	708.950	712.500	-0.5
2,4-D	8.497	8.412	8.612	682.320	705.000	-3.2
2,4-DB	10.261	10.179	10.379	625.670	712.500	-12.2
2,4-DCAA	7.357	7.269	7.469	709.420	750.000	-5.4
Dalapon	2.711	2.616	2.816	692.990	682.500	1.5
DICAMBA	7.548	7.461	7.661	710.810	705.000	0.8
DICHLORPROP	8.264	8.179	8.379	679.350	705.000	-3.6
Dinoseb	11.482	11.402	11.602	690.760	705.000	-2.0

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q2259 SAS No.: Q2259 SDG NO.: Q2259

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/04/2025 06/04/2025

Client Sample No.: CCAL04 Date Analyzed: 06/11/2025

Lab Sample No.: HSTDCCC750 Data File : PS030630.D Time Analyzed: 15:32

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.357	10.268	10.468	722.370	712.500	1.4
2,4,5-TP(Silvex)	9.931	9.842	10.042	724.770	712.500	1.7
2,4-D	9.023	8.934	9.134	714.950	705.000	1.4
2,4-DB	10.924	10.836	11.036	701.440	712.500	-1.6
2,4-DCAA	7.765	7.674	7.874	726.420	750.000	-3.1
Dalapon	2.707	2.610	2.810	673.700	682.500	-1.3
DICAMBA	7.968	7.877	8.077	715.390	705.000	1.5
DICHLORPROP	8.688	8.598	8.798	694.620	705.000	-1.5
Dinoseb	11.310	11.221	11.421	720.260	705.000	2.2