

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q2259 SAS No.: Q2259 SDG NO.: Q2259

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/16/2025 06/16/2025

Client Sample No.: CCAL05 Date Analyzed: 06/16/2025

Lab Sample No.: HSTDCCC750 Data File : PS030688.D Time Analyzed: 13:11

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.672	9.571	9.771	690.770	712.500	-3.0
2,4,5-TP(Silvex)	9.378	9.277	9.477	686.350	712.500	-3.7
2,4-D	8.488	8.388	8.588	663.780	705.000	-5.8
2,4-DB	10.251	10.150	10.350	688.140	712.500	-3.4
2,4-DCAA	7.350	7.250	7.450	686.030	750.000	-8.5
Dalapon	2.708	2.609	2.809	644.390	682.500	-5.6
DICAMBA	7.541	7.442	7.642	670.460	705.000	-4.9
DICHLORPROP	8.256	8.156	8.356	648.520	705.000	-8.0
Dinoseb	11.471	11.370	11.570	677.450	705.000	-3.9

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q2259 SAS No.: Q2259 SDG NO.: Q2259

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/16/2025 06/16/2025

Client Sample No.: CCAL05 Date Analyzed: 06/16/2025

Lab Sample No.: HSTDCCC750 Data File : PS030688.D Time Analyzed: 13:11

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.358	10.258	10.458	679.470	712.500	-4.6
2,4,5-TP(Silvex)	9.932	9.832	10.032	681.170	712.500	-4.4
2,4-D	9.025	8.926	9.126	661.470	705.000	-6.2
2,4-DB	10.925	10.825	11.025	666.500	712.500	-6.5
2,4-DCAA	7.768	7.668	7.868	699.780	750.000	-6.7
Dalapon	2.709	2.609	2.809	641.890	682.500	-6.0
DICAMBA	7.971	7.871	8.071	671.470	705.000	-4.8
DICHLORPROP	8.689	8.590	8.790	664.040	705.000	-5.8
Dinoseb	11.310	11.211	11.411	672.250	705.000	-4.6