

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q2259 SAS No.: Q2259 SDG NO.: Q2259

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/16/2025 06/16/2025

Client Sample No.: CCAL06 Date Analyzed: 06/16/2025

Lab Sample No.: HSTDCCC750 Data File : PS030698.D Time Analyzed: 17:25

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.672	9.571	9.771	782.520	712.500	9.8
2,4,5-TP(Silvex)	9.378	9.277	9.477	724.580	712.500	1.7
2,4-D	8.488	8.388	8.588	728.580	705.000	3.3
2,4-DB	10.250	10.150	10.350	803.270	712.500	12.7
2,4-DCAA	7.350	7.250	7.450	712.270	750.000	-5.0
Dalapon	2.708	2.609	2.809	653.700	682.500	-4.2
DICAMBA	7.541	7.442	7.642	685.590	705.000	-2.8
DICHLORPROP	8.256	8.156	8.356	683.810	705.000	-3.0
Dinoseb	11.470	11.370	11.570	726.650	705.000	3.1

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q2259 SAS No.: Q2259 SDG NO.: Q2259

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/16/2025 06/16/2025

Client Sample No.: CCAL06 Date Analyzed: 06/16/2025

Lab Sample No.: HSTDCCC750 Data File : PS030698.D Time Analyzed: 17:25

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.357	10.258	10.458	688.690	712.500	-3.3
2,4,5-TP(Silvex)	9.931	9.832	10.032	668.480	712.500	-6.2
2,4-D	9.025	8.926	9.126	667.720	705.000	-5.3
2,4-DB	10.925	10.825	11.025	666.720	712.500	-6.4
2,4-DCAA	7.768	7.668	7.868	706.810	750.000	-5.8
Dalapon	2.710	2.609	2.809	646.280	682.500	-5.3
DICAMBA	7.971	7.871	8.071	675.260	705.000	-4.2
DICHLORPROP	8.690	8.590	8.790	657.780	705.000	-6.7
Dinoseb	11.310	11.211	11.411	673.760	705.000	-4.4