

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q2259 SAS No.: Q2259 SDG NO.: Q2259

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/18/2025 06/18/2025

Client Sample No.: CCAL07 Date Analyzed: 06/19/2025

Lab Sample No.: HSTDCCC750 Data File : PS030753.D Time Analyzed: 09:25

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.668	9.568	9.768	726.800	712.500	2.0
2,4,5-TP(Silvex)	9.373	9.275	9.475	724.080	712.500	1.6
2,4-D	8.485	8.386	8.586	682.200	705.000	-3.2
2,4-DB	10.246	10.147	10.347	762.020	712.500	7.0
2,4-DCAA	7.348	7.249	7.449	757.570	750.000	1.0
Dalapon	2.713	2.611	2.811	696.660	682.500	2.1
DICAMBA	7.539	7.439	7.639	737.020	705.000	4.5
DICHLORPROP	8.253	8.153	8.353	673.640	705.000	-4.4
Dinoseb	11.465	11.366	11.566	715.310	705.000	1.5

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q2259 SAS No.: Q2259 SDG NO.: Q2259

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/18/2025 06/18/2025

Client Sample No.: CCAL07 Date Analyzed: 06/19/2025

Lab Sample No.: HSTDCCC750 Data File : PS030753.D Time Analyzed: 09:25

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.359	10.259	10.459	727.590	712.500	2.1
2,4,5-TP(Silvex)	9.933	9.833	10.033	717.060	712.500	0.6
2,4-D	9.027	8.927	9.127	684.160	705.000	-3.0
2,4-DB	10.926	10.827	11.027	712.330	712.500	0.0
2,4-DCAA	7.769	7.670	7.870	750.610	750.000	0.1
Dalapon	2.711	2.609	2.809	678.900	682.500	-0.5
DICAMBA	7.972	7.872	8.072	721.350	705.000	2.3
DICHLORPROP	8.691	8.591	8.791	686.560	705.000	-2.6
Dinoseb	11.312	11.211	11.411	712.700	705.000	1.1