

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q2259 SAS No.: Q2259 SDG NO.: Q2259

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 06/18/2025 06/18/2025

Client Sample No.: CCAL08 Date Analyzed: 06/19/2025

Lab Sample No.: HSTDCCC750 Data File : PS030757.D Time Analyzed: 11:06

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.667	9.568	9.768	747.680	712.500	4.9
2,4,5-TP(Silvex)	9.373	9.275	9.475	731.860	712.500	2.7
2,4-D	8.485	8.386	8.586	696.920	705.000	-1.1
2,4-DB	10.245	10.147	10.347	778.790	712.500	9.3
2,4-DCAA	7.348	7.249	7.449	769.350	750.000	2.6
Dalapon	2.713	2.611	2.811	695.430	682.500	1.9
DICAMBA	7.539	7.439	7.639	735.820	705.000	4.4
DICHLORPROP	8.252	8.153	8.353	679.830	705.000	-3.6
Dinoseb	11.465	11.366	11.566	710.420	705.000	0.8

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q2259 SAS No.: Q2259 SDG NO.: Q2259

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 06/18/2025 06/18/2025

Client Sample No.: CCAL08 Date Analyzed: 06/19/2025

Lab Sample No.: HSTDCCC750 Data File : PS030757.D Time Analyzed: 11:06

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.360	10.259	10.459	730.380	712.500	2.5
2,4,5-TP(Silvex)	9.934	9.833	10.033	724.230	712.500	1.6
2,4-D	9.027	8.927	9.127	711.340	705.000	0.9
2,4-DB	10.927	10.827	11.027	719.410	712.500	1.0
2,4-DCAA	7.771	7.670	7.870	749.750	750.000	0.0
Dalapon	2.713	2.609	2.809	675.650	682.500	-1.0
DICAMBA	7.973	7.872	8.072	720.320	705.000	2.2
DICHLORPROP	8.692	8.591	8.791	680.890	705.000	-3.4
Dinoseb	11.313	11.211	11.411	703.490	705.000	-0.2