

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS060425

Review By	Abdul	Review On	6/5/2025 8:33:42 AM
Supervise By	mohammad	Supervise On	6/6/2025 2:02:33 AM
SubDirectory	PS060425	HP Acquire Method	HP Processing Method ps060425 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560		
CCC Internal Standard/PEM	PP24559		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24562		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PS030474.D	04 Jun 2025 10:31	AR\AJ	Ok
2	I.BLK	PS030475.D	04 Jun 2025 10:55	AR\AJ	Ok
3	HSTDICC200	PS030476.D	04 Jun 2025 11:19	AR\AJ	Ok,M
4	HSTDICC500	PS030477.D	04 Jun 2025 11:43	AR\AJ	Ok,M
5	HSTDICC750	PS030478.D	04 Jun 2025 12:07	AR\AJ	Ok,M
6	HSTDICC1000	PS030479.D	04 Jun 2025 12:31	AR\AJ	Ok,M
7	HSTDICC1500	PS030480.D	04 Jun 2025 12:55	AR\AJ	Ok,M
8	HSTDICV750	PS030481.D	04 Jun 2025 13:35	AR\AJ	Ok,M
9	I.BLK	PS030482.D	04 Jun 2025 13:59	AR\AJ	Ok
10	HSTDCCC750	PS030483.D	04 Jun 2025 14:23	AR\AJ	Ok,M
11	PB168207BL	PS030484.D	04 Jun 2025 14:47	AR\AJ	Ok
12	PB168207BS	PS030485.D	04 Jun 2025 15:12	AR\AJ	Ok,M
13	Q1872-17	PS030486.D	04 Jun 2025 15:36	AR\AJ	Dilution
14	Q1872-17DL	PS030487.D	04 Jun 2025 16:00	AR\AJ	Ok,M
15	I.BLK	PS030488.D	04 Jun 2025 16:26	AR\AJ	Ok
16	HSTDCCC750	PS030489.D	04 Jun 2025 16:50	AR\AJ	Ok,M
17	Q2159-01	PS030490.D	04 Jun 2025 17:14	AR\AJ	Ok
18	Q2160-01	PS030491.D	04 Jun 2025 17:38	AR\AJ	Ok
19	Q2160-05	PS030492.D	04 Jun 2025 18:02	AR\AJ	Ok
20	Q2173-06	PS030493.D	04 Jun 2025 18:26	AR\AJ	Ok,M
21	Q2173-12	PS030494.D	04 Jun 2025 18:51	AR\AJ	Ok,M

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS060425

Review By	Abdul	Review On	6/5/2025 8:33:42 AM
Supervise By	mohammad	Supervise On	6/6/2025 2:02:33 AM
SubDirectory	PS060425	HP Acquire Method	HP Processing Method ps060425 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560		
CCC Internal Standard/PEM	PP24559		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24562		

22	Q2173-18	PS030495.D	04 Jun 2025 19:15	AR\AJ	Ok,M
23	I.BLK	PS030496.D	04 Jun 2025 19:39	AR\AJ	Ok
24	HSTDCCC750	PS030497.D	04 Jun 2025 20:03	AR\AJ	Ok,M
25	Q2173-06MS	PS030498.D	04 Jun 2025 20:51	AR\AJ	Ok,M
26	Q2173-06MSD	PS030499.D	04 Jun 2025 21:15	AR\AJ	Ok,M
27	Q2160-05MS	PS030500.D	04 Jun 2025 21:40	AR\AJ	Ok,M
28	Q2160-05MSD	PS030501.D	04 Jun 2025 22:04	AR\AJ	Ok,M
29	Q2172-01	PS030502.D	04 Jun 2025 22:28	AR\AJ	Ok
30	Q2185-01	PS030503.D	04 Jun 2025 22:52	AR\AJ	Ok,M
31	Q2185-05	PS030504.D	04 Jun 2025 23:16	AR\AJ	Ok
32	I.BLK	PS030505.D	04 Jun 2025 23:40	AR\AJ	Ok
33	HSTDCCC750	PS030506.D	05 Jun 2025 00:04	AR\AJ	Ok,M
34	Q2177-03	PS030507.D	05 Jun 2025 00:52	AR\AJ	Ok,M
35	Q2177-05	PS030508.D	05 Jun 2025 01:16	AR\AJ	Ok
36	Q2177-07	PS030509.D	05 Jun 2025 01:40	AR\AJ	Ok,M
37	PB168263BL	PS030510.D	05 Jun 2025 02:04	AR\AJ	Ok
38	PB168263BS	PS030511.D	05 Jun 2025 02:28	AR\AJ	Ok,M
39	PB168224TB	PS030512.D	05 Jun 2025 02:53	AR\AJ	Ok
40	PB168254BL	PS030513.D	05 Jun 2025 03:17	AR\AJ	Ok
41	PB168254BS	PS030514.D	05 Jun 2025 03:41	AR\AJ	Ok,M
42	I.BLK	PS030515.D	05 Jun 2025 04:05	AR\AJ	Ok
43	HSTDCCC750	PS030516.D	05 Jun 2025 04:29	AR\AJ	Ok,M
44	Q2176-01	PS030517.D	05 Jun 2025 05:17	AR\AJ	Ok

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS060425

Review By	Abdul	Review On	6/5/2025 8:33:42 AM
Supervise By	mohammad	Supervise On	6/6/2025 2:02:33 AM
SubDirectory	PS060425	HP Acquire Method	HP Processing Method ps060425 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560		
CCC Internal Standard/PEM	PP24559		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24562		

45	Q2176-02	PS030518.D	05 Jun 2025 05:41	AR\AJ	Ok,M
46	Q2176-03	PS030519.D	05 Jun 2025 06:05	AR\AJ	Ok
47	Q2176-04	PS030520.D	05 Jun 2025 06:29	AR\AJ	Not Ok
48	Q2176-05	PS030521.D	05 Jun 2025 06:53	AR\AJ	Ok
49	Q2176-06	PS030522.D	05 Jun 2025 07:17	AR\AJ	Ok,M
50	Q2176-07	PS030523.D	05 Jun 2025 07:41	AR\AJ	Ok,M
51	Q2176-08	PS030524.D	05 Jun 2025 08:05	AR\AJ	Ok
52	I.BLK	PS030525.D	05 Jun 2025 08:29	AR\AJ	Ok
53	HSTDCCC750	PS030526.D	05 Jun 2025 08:53	AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS061025

Review By	Abdul	Review On	6/11/2025 7:47:02 AM
Supervise By	mohammad	Supervise On	6/12/2025 3:33:36 AM
SubDirectory	PS061025	HP Acquire Method	HP Processing Method ps060425 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560		
CCC Internal Standard/PEM	PP24559		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24562		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PS030592.D	10 Jun 2025 11:39	AR\AJ	Ok
2	I.BLK	PS030593.D	10 Jun 2025 12:03	AR\AJ	Ok
3	HSTDCCC750	PS030594.D	10 Jun 2025 12:27	AR\AJ	Ok,M
4	Q2206-01	PS030595.D	10 Jun 2025 13:40	AR\AJ	Ok,M
5	Q2240-05	PS030596.D	10 Jun 2025 15:17	AR\AJ	ReRun
6	Q2240-09	PS030597.D	10 Jun 2025 15:48	AR\AJ	Not Ok
7	PB168375BL	PS030598.D	10 Jun 2025 16:48	AR\AJ	Not Ok
8	PB168375BS	PS030599.D	10 Jun 2025 17:12	AR\AJ	Not Ok
9	Q2246-01	PS030600.D	10 Jun 2025 17:36	AR\AJ	Ok,M
10	Q2259-01	PS030601.D	10 Jun 2025 18:00	AR\AJ	Ok,M
11	Q2259-03	PS030602.D	10 Jun 2025 18:24	AR\AJ	Ok
12	Q2260-01	PS030603.D	10 Jun 2025 18:48	AR\AJ	Ok,M
13	Q2150-06RE	PS030604.D	10 Jun 2025 19:12	AR\AJ	Not Ok
14	I.BLK	PS030605.D	10 Jun 2025 19:37	AR\AJ	Ok
15	HSTDCCC750	PS030606.D	10 Jun 2025 20:01	AR\AJ	Ok,M
16	Q2265-01	PS030607.D	10 Jun 2025 21:13	AR\AJ	Ok
17	Q2266-01	PS030608.D	10 Jun 2025 21:37	AR\AJ	Not Ok
18	Q2266-01MS	PS030609.D	10 Jun 2025 22:01	AR\AJ	Not Ok
19	Q2266-01MSD	PS030610.D	10 Jun 2025 22:25	AR\AJ	Not Ok
20	Q2266-05	PS030611.D	10 Jun 2025 22:49	AR\AJ	Ok
21	Q2271-01	PS030612.D	10 Jun 2025 23:13	AR\AJ	Ok

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS061025

Review By	Abdul	Review On	6/11/2025 7:47:02 AM
Supervise By	mohammad	Supervise On	6/12/2025 3:33:36 AM
SubDirectory	PS061025	HP Acquire Method	HP Processing Method ps060425 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560		
CCC Internal Standard/PEM	PP24559		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24562		

22	Q2272-01	PS030613.D	10 Jun 2025 23:37	AR\AJ	Ok
23	Q2273-01	PS030614.D	11 Jun 2025 00:02	AR\AJ	Ok
24	Q2273-05	PS030615.D	11 Jun 2025 00:26	AR\AJ	Ok
25	Q2274-01	PS030616.D	11 Jun 2025 00:50	AR\AJ	Ok
26	I.BLK	PS030617.D	11 Jun 2025 01:14	AR\AJ	Ok
27	HSTDCCC750	PS030618.D	11 Jun 2025 03:14	AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS061125

Review By	Abdul	Review On	6/12/2025 8:12:56 AM
Supervise By	mohammad	Supervise On	6/16/2025 2:04:29 AM
SubDirectory	PS061125	HP Acquire Method	HP Processing Method ps060425 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560		
CCC Internal Standard/PEM	PP24559		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24562		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PS030619.D	11 Jun 2025 08:22	AR\AJ	Ok
2	I.BLK	PS030620.D	11 Jun 2025 08:46	AR\AJ	Ok
3	HSTDCCC750	PS030621.D	11 Jun 2025 09:10	AR\AJ	Ok,M
4	Q2240-05	PS030622.D	11 Jun 2025 09:58	AR\AJ	Ok
5	Q2240-09	PS030623.D	11 Jun 2025 10:28	AR\AJ	Not Ok
6	Q2150-06	PS030624.D	11 Jun 2025 10:52	AR\AJ	Ok
7	Q2266-01	PS030625.D	11 Jun 2025 11:27	AR\AJ	Ok
8	Q2266-01MS	PS030626.D	11 Jun 2025 12:31	AR\AJ	Not Ok
9	Q2266-01MSD	PS030627.D	11 Jun 2025 12:55	AR\AJ	Not Ok
10	PB168375BS	PS030628.D	11 Jun 2025 14:06	AR\AJ	Ok,M
11	I.BLK	PS030629.D	11 Jun 2025 14:30	AR\AJ	Ok
12	HSTDCCC750	PS030630.D	11 Jun 2025 15:32	AR\AJ	Ok,M
13	Q2240-01RE	PS030631.D	11 Jun 2025 15:56	AR\AJ	Not Ok
14	I.BLK	PS030632.D	11 Jun 2025 16:23	AR\AJ	Ok
15	HSTDCCC750	PS030633.D	11 Jun 2025 16:48	AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS061625

Review By	Abdul	Review On	6/17/2025 8:57:56 AM
Supervise By	mohammad	Supervise On	6/18/2025 1:49:41 AM
SubDirectory	PS061625	HP Acquire Method	HP Processing Method ps061625 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560		
CCC Internal Standard/PEM	PP24559		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24562		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PS030679.D	16 Jun 2025 09:09	AR\AJ	Ok
2	I.BLK	PS030680.D	16 Jun 2025 09:32	AR\AJ	Ok
3	HSTDICC200	PS030681.D	16 Jun 2025 09:56	AR\AJ	Ok,M
4	HSTDICC500	PS030682.D	16 Jun 2025 10:20	AR\AJ	Ok,M
5	HSTDICC750	PS030683.D	16 Jun 2025 10:44	AR\AJ	Ok,M
6	HSTDICC1000	PS030684.D	16 Jun 2025 11:09	AR\AJ	Ok,M
7	HSTDICC1500	PS030685.D	16 Jun 2025 11:33	AR\AJ	Ok,M
8	HSTDICV750	PS030686.D	16 Jun 2025 11:57	AR\AJ	Ok,M
9	I.BLK	PS030687.D	16 Jun 2025 12:23	AR\AJ	Ok
10	HSTDCCC750	PS030688.D	16 Jun 2025 13:11	AR\AJ	Ok,M
11	Q2287-01	PS030689.D	16 Jun 2025 13:35	AR\AJ	Ok
12	PB168465BL	PS030690.D	16 Jun 2025 13:59	AR\AJ	Ok
13	PB168465BS	PS030691.D	16 Jun 2025 14:23	AR\AJ	Ok,M
14	PB168375BL	PS030692.D	16 Jun 2025 14:53	AR\AJ	Ok
15	Q2286-03RE	PS030693.D	16 Jun 2025 15:17	AR\AJ	Confirms
16	Q2296-01	PS030694.D	16 Jun 2025 15:49	AR\AJ	Ok
17	Q2296-05	PS030695.D	16 Jun 2025 16:13	AR\AJ	Ok
18	Q2296-09	PS030696.D	16 Jun 2025 16:37	AR\AJ	Ok
19	I.BLK	PS030697.D	16 Jun 2025 17:01	AR\AJ	Ok
20	HSTDCCC750	PS030698.D	16 Jun 2025 17:25	AR\AJ	Ok,M
21	PB168495BL	PS030699.D	16 Jun 2025 18:13	AR\AJ	Ok

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QCBatch ID # PS061625

Review By	Abdul	Review On	6/17/2025 8:57:56 AM
Supervise By	mohammad	Supervise On	6/18/2025 1:49:41 AM
SubDirectory	PS061625	HP Acquire Method	HP Processing Method ps061625 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560		
CCC Internal Standard/PEM	PP24559		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24562		

22	PB168495BS	PS030700.D	16 Jun 2025 18:37	AR\AJ	Ok,M
23	PB168428TB	PS030701.D	16 Jun 2025 19:01	AR\AJ	Ok
24	Q2301-03	PS030702.D	16 Jun 2025 19:26	AR\AJ	Ok
25	Q2301-03MS	PS030703.D	16 Jun 2025 19:50	AR\AJ	Ok,M
26	Q2301-03MSD	PS030704.D	16 Jun 2025 20:14	AR\AJ	Ok,M
27	I.BLK	PS030705.D	16 Jun 2025 20:38	AR\AJ	Ok
28	HSTDCCC750	PS030706.D	16 Jun 2025 21:02	AR\AJ	Ok,M
29	Q2296-13	PS030707.D	16 Jun 2025 21:50	AR\AJ	Ok,M
30	Q2296-13MS	PS030708.D	16 Jun 2025 22:14	AR\AJ	Ok,M
31	Q2296-13MSD	PS030709.D	16 Jun 2025 22:38	AR\AJ	Ok,M
32	Q2296-17	PS030710.D	16 Jun 2025 23:02	AR\AJ	Ok
33	Q2296-21	PS030711.D	16 Jun 2025 23:26	AR\AJ	Ok,M
34	Q2297-01	PS030712.D	16 Jun 2025 23:50	AR\AJ	Ok
35	Q2298-01	PS030713.D	17 Jun 2025 00:14	AR\AJ	Ok
36	Q2310-01	PS030714.D	17 Jun 2025 00:38	AR\AJ	Ok,M
37	I.BLK	PS030715.D	17 Jun 2025 01:02	AR\AJ	Ok
38	HSTDCCC750	PS030716.D	17 Jun 2025 01:26	AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS061825

Review By	Abdul	Review On	6/19/2025 8:25:06 AM
Supervise By	mohammad	Supervise On	6/20/2025 8:27:44 AM
SubDirectory	PS061825	HP Acquire Method	HP Processing Method PS061825 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560		
CCC Internal Standard/PEM	PP24559		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24562		

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PS030737.D	18 Jun 2025 10:37	AR\AJ	Ok
2	I.BLK	PS030738.D	18 Jun 2025 11:01	AR\AJ	Ok
3	HSTDICC200	PS030739.D	18 Jun 2025 11:25	AR\AJ	Ok,M
4	HSTDICC500	PS030740.D	18 Jun 2025 11:49	AR\AJ	Ok
5	HSTDICC750	PS030741.D	18 Jun 2025 12:13	AR\AJ	Ok,M
6	HSTDICC1000	PS030742.D	18 Jun 2025 12:37	AR\AJ	Ok,M
7	HSTDICC1500	PS030743.D	18 Jun 2025 13:29	AR\AJ	Ok,M
8	HSTDICV750	PS030744.D	18 Jun 2025 14:18	AR\AJ	Ok,M
9	I.BLK	PS030745.D	18 Jun 2025 14:44	AR\AJ	Ok
10	HSTDCCC750	PS030746.D	18 Jun 2025 15:08	AR\AJ	Ok,M
11	Q2312-01MS	PS030747.D	18 Jun 2025 17:07	AR\AJ	Ok,M
12	Q2312-01MSD	PS030748.D	18 Jun 2025 17:31	AR\AJ	Ok,M
13	I.BLK	PS030749.D	18 Jun 2025 17:55	AR\AJ	Ok
14	HSTDCCC750	PS030750.D	18 Jun 2025 18:19	AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS061925

Review By	Abdul	Review On	6/20/2025 8:24:30 AM
Supervise By	mohammad	Supervise On	6/20/2025 8:27:18 AM
SubDirectory	PS061925	HP Acquire Method	HP Processing Method PS061825 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560		
CCC Internal Standard/PEM	PP24559		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24562		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PS030751.D	19 Jun 2025 08:37	AR\AJ	Ok
2	I.BLK	PS030752.D	19 Jun 2025 09:01	AR\AJ	Ok
3	HSTDCCC750	PS030753.D	19 Jun 2025 09:25	AR\AJ	Ok,M
4	Q2266-01MS	PS030754.D	19 Jun 2025 09:50	AR\AJ	Ok,M
5	Q2266-01MSD	PS030755.D	19 Jun 2025 10:18	AR\AJ	Ok,M
6	I.BLK	PS030756.D	19 Jun 2025 10:42	AR\AJ	Ok
7	HSTDCCC750	PS030757.D	19 Jun 2025 11:06	AR\AJ	Ok,M
8	Q2126-01	PS030758.D	19 Jun 2025 17:16	AR\AJ	Not Ok
9	Q2126-07	PS030759.D	19 Jun 2025 17:40	AR\AJ	Not Ok
10	Q2126-01	PS030760.D	19 Jun 2025 18:04	AR\AJ	Ok,M
11	Q2126-07	PS030761.D	19 Jun 2025 18:29	AR\AJ	Ok,M
12	Q2126-02	PS030762.D	19 Jun 2025 18:53	AR\AJ	Ok,M
13	Q2126-08	PS030763.D	19 Jun 2025 19:17	AR\AJ	Ok,M
14	I.BLK	PS030764.D	19 Jun 2025 19:41	AR\AJ	Ok
15	HSTDCCC750	PS030765.D	19 Jun 2025 20:05	AR\AJ	Ok,M

M : Manual Integration