

Report of Analysis

Client:	Nobis Group				Date Collected:	06/18/2	5		
Project: Raymark Supe		rfund Site			Date Received:	06/18/2	06/18/25		
Client Sample ID:	PIBLK-PS030	PIBLK-PS030738.D				Q2259			
Lab Sample ID:	I BLK-PS0307	LK-PS030738.D			Matrix:	WATER	WATER		
•		56.0						1	
Analytical Method	: 8151A				% Solid:	0	Decanted	1:	
Sample Wt/Vol:	1000 Un	iits: mL			Final Vol:	10000	uL		
Soil Aliquot Vol:		uL	uL		Test:	Herbicide Group1			
Extraction Type:					Injection Volume :				
GPC Factor :	1.0	PH :							
Prep Method :	SW3510C								
Trop Wethod .	51135100								
File ID/Qc Batch:	ile ID/Qc Batch: Dilution:		Prep Date		Date Analyzed		Prep Batch ID		
PS030738.D	1				06/18/25		PS061825		
CAS Number	Parameter	Conc.	Qualifier	MDL		LOD LOQ	LOD LOQ / CRQL Uni		
TARGETS									
1918-00-9	DICAMBA	0.0015	U	0.00065		0.0015	0.0020	mg/L	
75-99-0	DALAPON	0.0015	U	0.00098		0.0015	0.0020	mg/L	
120-36-5	DICHLORPROP	0.0015	U	0.00076		0.0015	0.0020	mg/L	
94-75-7	2,4-D	0.0015	U	0.00092		0.0015	0.0020	mg/L	
93-72-1	2,4,5-TP (Silvex)	0.0015	U	0.00078		0.0015	0.0020	mg/L	
93-76-5	2,4,5-T	0.0015	U	0.00071		0.0015	0.0020	mg/L	
94-82-6	2,4-DB	0.0015	U	0.00065		0.0015	0.0020	mg/L	
88-85-7	DINOSEB	0.0015	U	0.00089		0.0015	0.0020	mg/L	
SURROGATES									

Comments:

U = Not Detected

LOQ = Limit of Quantitation

- MDL = Method Detection Limit
- LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates > 25% difference for detected

concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

- J = Estimated Value
- B = Analyte Found in Associated Method Blank
- N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration

was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit