

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS061125

Review By	Abdul	Review On	6/12/2025 8:12:56 AM
Supervise By	mohammad	Supervise On	6/16/2025 2:04:29 AM
SubDirectory	PS061125	HP Acquire Method	HP Processing Method ps060425 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560		
CCC	PP24559		
Internal Standard/PEM			
ICV/I.BLK	PP24562		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS030619.D	11 Jun 2025 08:22		AR\AJ	Ok
2	I.BLK	I.BLK	PS030620.D	11 Jun 2025 08:46		AR\AJ	Ok
3	HSTDCCC750	HSTDCCC750	PS030621.D	11 Jun 2025 09:10		AR\AJ	Ok,M
4	Q2240-05	TP-2	PS030622.D	11 Jun 2025 09:58		AR\AJ	Ok
5	Q2240-09	TP-1	PS030623.D	11 Jun 2025 10:28	wrong Injection	AR\AJ	Not Ok
6	Q2150-06	TP-50	PS030624.D	11 Jun 2025 10:52		AR\AJ	Ok
7	Q2266-01	WC-3	PS030625.D	11 Jun 2025 11:27		AR\AJ	Ok
8	Q2266-01MS	WC-3MS	PS030626.D	11 Jun 2025 12:31	some compound recovery fail	AR\AJ	Not Ok
9	Q2266-01MSD	WC-3MSD	PS030627.D	11 Jun 2025 12:55	some compound recovery fail	AR\AJ	Not Ok
10	PB168375BS	PB168375BS	PS030628.D	11 Jun 2025 14:06		AR\AJ	Ok,M
11	I.BLK	I.BLK	PS030629.D	11 Jun 2025 14:30		AR\AJ	Ok
12	HSTDCCC750	HSTDCCC750	PS030630.D	11 Jun 2025 15:32		AR\AJ	Ok,M
13	Q2240-01RE	TP-3RE	PS030631.D	11 Jun 2025 15:56	Already run	AR\AJ	Not Ok
14	I.BLK	I.BLK	PS030632.D	11 Jun 2025 16:23		AR\AJ	Ok
15	HSTDCCC750	HSTDCCC750	PS030633.D	11 Jun 2025 16:48		AR\AJ	Ok,M

M : Manual Integration