

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS061925

Review By	Abdul	Review On	6/20/2025 8:24:30 AM
Supervise By	mohammad	Supervise On	6/20/2025 8:27:18 AM
SubDirectory	PS061925	HP Acquire Method	HP Processing Method PS061825 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24553,PP24554,PP24556,PP24557,PP24558,PP24559,PP24560
CCC	PP24559
Internal Standard/PEM	
ICV/I.BLK	PP24562
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS030751.D	19 Jun 2025 08:37		AR\AJ	Ok
2	I.BLK	I.BLK	PS030752.D	19 Jun 2025 09:01		AR\AJ	Ok
3	HSTDCCC750	HSTDCCC750	PS030753.D	19 Jun 2025 09:25		AR\AJ	Ok,M
4	Q2266-01MS	WC-3MS	PS030754.D	19 Jun 2025 09:50	some compound recovery fail	AR\AJ	Ok,M
5	Q2266-01MSD	WC-3MSD	PS030755.D	19 Jun 2025 10:18	some compound recovery fail	AR\AJ	Ok,M
6	I.BLK	I.BLK	PS030756.D	19 Jun 2025 10:42		AR\AJ	Ok
7	HSTDCCC750	HSTDCCC750	PS030757.D	19 Jun 2025 11:06		AR\AJ	Ok,M
8	Q2126-01	LOD-MDL-SOIL-03-QT	PS030758.D	19 Jun 2025 17:16	Surrogate high in both column, LOD 100PPB	AR\AJ	Not Ok
9	Q2126-07	LOD-MDL-WATER-01-Q	PS030759.D	19 Jun 2025 17:40	Surrogate high in both column , LOD 100PPB	AR\AJ	Not Ok
10	Q2126-01	LOD-MDL-SOIL-03-QT	PS030760.D	19 Jun 2025 18:04	LOD 150 PPB	AR\AJ	Ok,M
11	Q2126-07	LOD-MDL-WATER-01-Q	PS030761.D	19 Jun 2025 18:29	LOD 150 PPB	AR\AJ	Ok,M
12	Q2126-02	LOQ-SOIL-02-QT2-202	PS030762.D	19 Jun 2025 18:53	LOQ 200 PPB	AR\AJ	Ok,M
13	Q2126-08	LOQ-WATER-02-QT2-2	PS030763.D	19 Jun 2025 19:17	LOQ 200 PPB	AR\AJ	Ok,M
14	I.BLK	I.BLK	PS030764.D	19 Jun 2025 19:41		AR\AJ	Ok
15	HSTDCCC750	HSTDCCC750	PS030765.D	19 Jun 2025 20:05		AR\AJ	Ok,M

M : Manual Integration