

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q2259 SAS No.: Q2259 SDG NO.: Q2259

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 06/09/2025

Lab Sample No.: AR1660CCC500 Data File : PP072732.D Time Analyzed: 09:22

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.650	5.552	5.752	466.900	500.000	-6.6
Aroclor-1016-2	5.671	5.574	5.774	500.040	500.000	0.0
Aroclor-1016-3	5.733	5.636	5.836	501.080	500.000	0.2
Aroclor-1016-4	5.831	5.734	5.934	527.290	500.000	5.5
Aroclor-1016-5	6.123	6.027	6.227	407.640	500.000	-18.5
Aroclor-1260-1	7.241	7.147	7.347	486.440	500.000	-2.7
Aroclor-1260-2	7.494	7.401	7.601	485.810	500.000	-2.8
Aroclor-1260-3	7.852	7.760	7.960	504.460	500.000	0.9
Aroclor-1260-4	8.077	7.985	8.185	501.920	500.000	0.4
Aroclor-1260-5	8.395	8.303	8.503	534.190	500.000	6.8
Decachlorobiphenyl	10.196	10.113	10.313	52.060	50.000	4.1
Tetrachloro-m-xylene	4.498	4.396	4.596	53.360	50.000	6.7

CALIBRATION VERIFICATION SUMMARY

Contract: NOBI03

Lab Code: CHEM Case No.: Q2259 SAS No.: Q2259 SDG NO.: Q2259

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 06/09/2025

Lab Sample No.: AR1660CCC500 Data File : PP072732.D Time Analyzed: 09:22

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.874	4.777	4.977	512.090	500.000	2.4
Aroclor-1016-2	4.892	4.795	4.995	526.060	500.000	5.2
Aroclor-1016-3	5.068	4.973	5.173	531.050	500.000	6.2
Aroclor-1016-4	5.110	5.015	5.215	536.050	500.000	7.2
Aroclor-1016-5	5.324	5.229	5.429	537.990	500.000	7.6
Aroclor-1260-1	6.355	6.264	6.464	537.850	500.000	7.6
Aroclor-1260-2	6.544	6.453	6.653	517.950	500.000	3.6
Aroclor-1260-3	6.696	6.606	6.806	542.980	500.000	8.6
Aroclor-1260-4	7.167	7.077	7.277	547.180	500.000	9.4
Aroclor-1260-5	7.409	7.319	7.519	551.870	500.000	10.4
Decachlorobiphenyl	8.806	8.723	8.923	55.450	50.000	10.9
Tetrachloro-m-xylene	3.793	3.691	3.891	53.430	50.000	6.9