

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJEC	T NUMBER:	bf061025					
SequenceID :	bf061025				NA	NO	YES
1. Chromatograms Labe	eled/Compounds Id	entified. (Field samples	and Method Blanks)		<u>✓</u>		
 2. GC/MS Tuning Specifications. DFTPP Meet Criteria Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ) 3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series 					✓		
					✓		
 GC/MS Calibration - analysis and continui for 600 series and 12 	ng calibration perfo	rmed within 24 hours of			√		
5. GC/MS Calibration M	Met:				✓		
a. Initial calibration I If not met, list those cor		ecoveries which fall out	side the acceptable range.		√		
	npounds and their r	ecoveries which fall out	side the acceptable range.	re list of the same			
 Blank Contamination 	-			is list of the sample			

a. B/N Fraction

d. Acid Fraction

7. Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges.	<u> </u>	
a. B/N Fraction		
d. Acid Fraction		
8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet CriteriaIf not met, list those compounds and their recoveries which fall outside the acceptable range.a. B/N Fraction	<u>✓</u>	
d. Acid Fraction		
9. Internal Standard Area/Retention Time Shift Meet Criteria Comments:	<u>✓</u>	
10. Extraction Holding Time Met If not met, list number of days exceeded for each sample:	<u> </u>	
11. Analysis Holding Time Met If not met, list number of days exceeded for each sample:	<u>√</u>	

ADDITIONAL COMMENTS:

Recovery of 3,3-Dichlorobenzidine is marginally biased low and few compounds are slightly biased high in the PB168351BS. The data will be used for hardcopies.



NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER:

SequenceID :

bf061025

bf061025

NA NO YES