

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2259

Client: Nobis Group

Analytical Method: 8270E

DataFile: BF142703.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB168351BS	Naphthalene	1700	1600	ug/Kg	94				35	123	
	2-Methylnaphthalene	1700	1700	ug/Kg	100				38	122	
	Acenaphthylene	1700	1600	ug/Kg	94				32	132	
	Acenaphthene	1700	1700	ug/Kg	100				40	123	
	Fluorene	1700	1500	ug/Kg	88				43	125	
	Phenanthrene	1700	1600	ug/Kg	94				50	121	
	Anthracene	1700	1600	ug/Kg	94				47	123	
	Fluoranthene	1700	1700	ug/Kg	100				50	127	
	Pyrene	1700	1800	ug/Kg	106				47	127	
	Benzo(a)anthracene	1700	1700	ug/Kg	100				49	126	
	Chrysene	1700	1600	ug/Kg	94				50	124	
	Benzo(b)fluoranthene	1700	2100	ug/Kg	124				45	132	
	Benzo(k)fluoranthene	1700	2000	ug/Kg	118				47	132	
	Benzo(a)pyrene	1700	1700	ug/Kg	100				45	129	
	Indeno(1,2,3-cd)pyrene	1700	2100	ug/Kg	124				45	133	
	Dibenz(a,h)anthracene	1700	2100	ug/Kg	124				45	134	
	Benzo(g,h,i)perylene	1700	2100	ug/Kg	124				43	134	