

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>NOBI03</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2259</u>
Lab File ID:	<u>VY022638.D</u>	Date Analyzed:	<u>06/10/2025</u>
Instrument ID:	<u>MSVOA_Y</u>	Time Analyzed:	<u>09:46</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)
		Heated Purge:	(Y/N) <u>Y</u>

	IS4 AREA #	RT #				
12 HOUR STD	106365	13.347				
UPPER LIMIT	212730	13.847				
LOWER LIMIT	53182.5	12.847				
EPA SAMPLE NO.						
OU4-PCS-TC-36-060525	119411	13.35				
OU4-PCS-TC-37-060525	99853	13.35				
VY0610SBL01	122259	13.35				
VY0610SBS01	109725	13.35				
VY0610SBSD01	107638	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.