

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: NOBI03
 Lab Code: CHEM Case No.: Q2259 SAS No.: Q2259 SDG NO.: Q2259
 Lab File ID: VY022652.D Date Analyzed: 06/11/2025
 Instrument ID: MSVOA_Y Time Analyzed: 09:59
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	106876	13.346				
UPPER LIMIT	213752	13.846				
LOWER LIMIT	53438	12.846				
EPA SAMPLE NO.						
OU4-PCS-TC-36-060525	66012	13.35				
OU4-PCS-TC-37-060525	86500	13.35				
VY0611SBL01	84338	13.35				
VY0611SBS01	105447	13.35				
VY0611SBSD01	102308	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.