

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2263

Client: AECOM Technical Services, Inc.

Analytical Method: 8270-Modified DataFile: BN037236.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB168391BS	1,4-Dioxane	0.4	0.39	ug/L	98				70	130	