

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2263 SAS No.: Q2263 SDG NO.: Q2263

EPA Sample No.: SSTDICCC0.4 Date Analyzed: 06/13/2025

Lab File ID: BN037227.D Time Analyzed: 14:46

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	1287	7.575	3210	10.36	1738	14.22
UPPER LIMIT	2574	8.075	6420	10.861	3476	14.724
LOWER LIMIT	643.5	7.075	1605	9.861	869	13.724
EPA SAMPLE NO.						
01 PB168391BS	1477	7.58	3518	10.35	1759	14.22
02 PB168391BSD	1340	7.58	3197	10.35	1517	14.22
03 PB168391BL	1036	7.58	2301	10.37	1224	14.23

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.