

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2263 SAS No.: Q2263 SDG NO.: Q2263

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 06/13/2025

Lab File ID: BN037240.D Time Analyzed: 23:55

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	1101	7.582	2581	10.35	1319	14.22
UPPER LIMIT	2202	8.082	5162	10.851	2638	14.724
LOWER LIMIT	550.5	7.082	1290.5	9.851	659.5	13.724
EPA SAMPLE NO.						
01 RW9-MW01D3-20250606	958	7.58	2256	10.36	1255	14.22
02 RW9-MW01S-20250606	932	7.58	2363	10.36	1257	14.22

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.