

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2263 SAS No.: Q2263 SDG NO.: Q2263

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 06/13/2025

Lab File ID: BN037240.D Time Analyzed: 23:55

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	2396	16.971	1787	21.171	1797	23.36
UPPER LIMIT	4792	17.471	3574	21.671	3594	23.86
LOWER LIMIT	1198	16.471	893.5	20.671	898.5	22.86
EPA SAMPLE NO.						
01 RW9-MW01D3-20250606	2156	16.98	1827	21.17	1841	23.36
02 RW9-MW01S-20250606	2205	16.98	1783	21.17	1754	23.36

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.