

LAB CHRONICLE

OrderID:	Q2263	OrderDate:	6/6/2025 12:42:00 PM
Client:	AECOM Technical Services, Inc.	Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258
Contact:	Eleanor Vivadou	Location:	D21,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2263-01	RW9-MW01D3-20250606	Water	VOCMS Group1	8260-Low	06/06/25		06/11/25	06/06/25
Q2263-02	RW9-MW01S-20250606	Water	VOCMS Group1	8260-Low	06/06/25		06/11/25	06/06/25



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q2263

Client: AECOM Technical Services, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID:	RW9-MW01D3-20250606								
Q2263-01	RW9-MW01D3-20250606	Water	Acetone	1.90	J	1.50	3.80	5.00	ug/L
			Total Voc :	1.90					
			Total Concentration:	1.90					



SAMPLE DATA

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	06/06/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	06/06/25	
Client Sample ID:	RW9-MW01D3-20250606		SDG No.:	Q2263	
Lab Sample ID:	Q2263-01		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086963.D	1		06/11/25 20:39	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	1.90	J	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	06/06/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	06/06/25	
Client Sample ID:	RW9-MW01D3-20250606		SDG No.:	Q2263	
Lab Sample ID:	Q2263-01		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086963.D	1		06/11/25 20:39	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	50.6		81 - 118		101%	SPK: 50
1868-53-7	Dibromofluoromethane	50.7		80 - 119		101%	SPK: 50
2037-26-5	Toluene-d8	51.9		89 - 112		104%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.0		85 - 114		100%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	256000	8.235				
540-36-3	1,4-Difluorobenzene	487000	9.106				
3114-55-4	Chlorobenzene-d5	434000	11.865				
3855-82-1	1,4-Dichlorobenzene-d4	207000	13.794				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	06/06/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	06/06/25	
Client Sample ID:	RW9-MW01D3-20250606		SDG No.:	Q2263	
Lab Sample ID:	Q2263-01		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086963.D	1		06/11/25 20:39	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	06/06/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	06/06/25	
Client Sample ID:	RW9-MW01S-20250606		SDG No.:	Q2263	
Lab Sample ID:	Q2263-02		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086964.D	1		06/11/25 21:01	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	06/06/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	06/06/25	
Client Sample ID:	RW9-MW01S-20250606		SDG No.:	Q2263	
Lab Sample ID:	Q2263-02		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086964.D	1		06/11/25 21:01	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.8		81 - 118		108%	SPK: 50
1868-53-7	Dibromofluoromethane	51.3		80 - 119		103%	SPK: 50
2037-26-5	Toluene-d8	51.7		89 - 112		103%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.0		85 - 114		100%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	212000	8.235				
540-36-3	1,4-Difluorobenzene	417000	9.106				
3114-55-4	Chlorobenzene-d5	370000	11.865				
3855-82-1	1,4-Dichlorobenzene-d4	174000	13.794				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	06/06/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	06/06/25	
Client Sample ID:	RW9-MW01S-20250606		SDG No.:	Q2263	
Lab Sample ID:	Q2263-02		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086964.D	1		06/11/25 21:01	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SDG No.: Q2263

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2263-01	RW9-MW01D3-20250606	1,2-Dichloroethane-d4	50	50.6	101	81	118
		Dibromofluoromethane	50	50.8	101	80	119
		Toluene-d8	50	51.9	104	89	112
		4-Bromofluorobenzene	50	50.0	100	85	114
Q2263-02	RW9-MW01S-20250606	1,2-Dichloroethane-d4	50	53.8	108	81	118
		Dibromofluoromethane	50	51.3	103	80	119
		Toluene-d8	50	51.7	103	89	112
		4-Bromofluorobenzene	50	50.0	100	85	114
VN0611WBL01	VN0611WBL01	1,2-Dichloroethane-d4	50	48.3	97	81	118
		Dibromofluoromethane	50	49.3	99	80	119
		Toluene-d8	50	51.7	103	89	112
		4-Bromofluorobenzene	50	49.4	99	85	114
VN0611WBS02	VN0611WBS02	1,2-Dichloroethane-d4	50	47.1	94	81	118
		Dibromofluoromethane	50	51.1	102	80	119
		Toluene-d8	50	48.7	97	89	112
		4-Bromofluorobenzene	50	50.0	100	85	114
VN0611WBSD0	VN0611WBSD02	1,2-Dichloroethane-d4	50	47.9	96	81	118
		Dibromofluoromethane	50	52.3	105	80	119
		Toluene-d8	50	49.5	99	89	112
		4-Bromofluorobenzene	50	50.1	100	85	114

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2263

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VN086944.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0611WBS02	Dichlorodifluoromethane	20	19.7	ug/L	99			32	152	
	Chloromethane	20	16.1	ug/L	81			50	139	
	Vinyl chloride	20	18.9	ug/L	95			58	137	
	Bromomethane	20	16.6	ug/L	83			53	141	
	Chloroethane	20	19.4	ug/L	97			60	138	
	Trichlorofluoromethane	20	19.2	ug/L	96			65	141	
	1,1,2-Trichlorotrifluoroethane	20	19.6	ug/L	98			70	136	
	1,1-Dichloroethene	20	19.2	ug/L	96			71	131	
	Acetone	100	99.0	ug/L	99			39	160	
	Carbon disulfide	20	17.8	ug/L	89			64	133	
	Methyl tert-butyl Ether	20	19.6	ug/L	98			71	124	
	Methyl Acetate	20	19.0	ug/L	95			56	136	
	Methylene Chloride	20	19.0	ug/L	95			74	124	
	trans-1,2-Dichloroethene	20	18.5	ug/L	93			75	124	
	1,1-Dichloroethane	20	19.4	ug/L	97			77	125	
	Cyclohexane	20	17.6	ug/L	88			71	130	
	2-Butanone	100	90.4	ug/L	90			56	143	
	Carbon Tetrachloride	20	19.2	ug/L	96			72	136	
	cis-1,2-Dichloroethene	20	19.5	ug/L	98			78	123	
	Bromochloromethane	20	19.9	ug/L	100			78	123	
	Chloroform	20	19.5	ug/L	98			79	124	
	1,1,1-Trichloroethane	20	19.0	ug/L	95			74	131	
	Methylcyclohexane	20	16.9	ug/L	85			72	132	
	Benzene	20	19.3	ug/L	97			79	120	
	1,2-Dichloroethane	20	19.5	ug/L	98			73	128	
	Trichloroethene	20	19.9	ug/L	100			79	123	
	1,2-Dichloropropane	20	19.6	ug/L	98			78	122	
	Bromodichloromethane	20	19.6	ug/L	98			79	125	
	4-Methyl-2-Pentanone	100	98.0	ug/L	98			67	130	
	Toluene	20	19.5	ug/L	98			80	121	
	t-1,3-Dichloropropene	20	19.9	ug/L	100			73	127	
	cis-1,3-Dichloropropene	20	20.1	ug/L	101			75	124	
	1,1,2-Trichloroethane	20	20.4	ug/L	102			80	119	
	2-Hexanone	100	89.2	ug/L	89			57	139	
	Dibromochloromethane	20	20.5	ug/L	103			74	126	
	1,2-Dibromoethane	20	19.7	ug/L	99			77	121	
	Tetrachloroethene	20	19.2	ug/L	96			74	129	
	Chlorobenzene	20	20.0	ug/L	100			82	118	
	Ethyl Benzene	20	19.6	ug/L	98			79	121	
	m/p-Xylenes	40	39.6	ug/L	99			80	121	
	o-Xylene	20	20.2	ug/L	101			78	122	
	Styrene	20	20.1	ug/L	101			78	123	
	Bromoform	20	21.0	ug/L	105			66	130	
	Isopropylbenzene	20	19.3	ug/L	97			72	131	
	1,1,2,2-Tetrachloroethane	20	21.0	ug/L	105			71	121	
	1,3-Dichlorobenzene	20	20.3	ug/L	102			80	119	
	1,4-Dichlorobenzene	20	20.4	ug/L	102			79	118	
	1,2-Dichlorobenzene	20	20.0	ug/L	100			80	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2263

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VN086944.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VN0611WBS02	1,2-Dibromo-3-Chloropropane	20	19.2	ug/L	96			62	128	
	1,2,4-Trichlorobenzene	20	18.4	ug/L	92			69	130	
	1,2,3-Trichlorobenzene	20	17.8	ug/L	89			69	129	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2263

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VN086945.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0611WBSD02	Dichlorodifluoromethane	20	19.1	ug/L	96	3		32	152	20
	Chloromethane	20	15.5	ug/L	78	4		50	139	20
	Vinyl chloride	20	18.7	ug/L	94	1		58	137	20
	Bromomethane	20	16.2	ug/L	81	2		53	141	20
	Chloroethane	20	18.9	ug/L	95	2		60	138	20
	Trichlorofluoromethane	20	19.0	ug/L	95	1		65	141	20
	1,1,2-Trichlorotrifluoroethane	20	18.7	ug/L	94	4		70	136	20
	1,1-Dichloroethene	20	19.1	ug/L	96	0		71	131	20
	Acetone	100	92.4	ug/L	92	7		39	160	20
	Carbon disulfide	20	17.2	ug/L	86	3		64	133	20
	Methyl tert-butyl Ether	20	20.2	ug/L	101	3		71	124	20
	Methyl Acetate	20	20.1	ug/L	101	6		56	136	20
	Methylene Chloride	20	19.1	ug/L	96	1		74	124	20
	trans-1,2-Dichloroethene	20	18.6	ug/L	93	0		75	124	20
	1,1-Dichloroethane	20	19.3	ug/L	97	0		77	125	20
	Cyclohexane	20	17.2	ug/L	86	2		71	130	20
	2-Butanone	100	94.9	ug/L	95	5		56	143	20
	Carbon Tetrachloride	20	19.5	ug/L	98	2		72	136	20
	cis-1,2-Dichloroethene	20	19.6	ug/L	98	0		78	123	20
	Bromochloromethane	20	21.1	ug/L	106	6		78	123	20
	Chloroform	20	19.4	ug/L	97	1		79	124	20
	1,1,1-Trichloroethane	20	18.8	ug/L	94	1		74	131	20
	Methylcyclohexane	20	16.6	ug/L	83	2		72	132	20
	Benzene	20	19.7	ug/L	99	2		79	120	20
	1,2-Dichloroethane	20	20.3	ug/L	102	4		73	128	20
	Trichloroethene	20	20.1	ug/L	101	1		79	123	20
	1,2-Dichloropropane	20	20.0	ug/L	100	2		78	122	20
	Bromodichloromethane	20	20.7	ug/L	104	6		79	125	20
	4-Methyl-2-Pentanone	100	100	ug/L	100	2		67	130	20
	Toluene	20	19.9	ug/L	100	2		80	121	20
	t-1,3-Dichloropropene	20	20.7	ug/L	104	4		73	127	20
	cis-1,3-Dichloropropene	20	20.7	ug/L	104	3		75	124	20
	1,1,2-Trichloroethane	20	20.9	ug/L	104	2		80	119	20
	2-Hexanone	100	93.8	ug/L	94	5		57	139	20
	Dibromochloromethane	20	21.2	ug/L	106	3		74	126	20
	1,2-Dibromoethane	20	20.7	ug/L	104	5		77	121	20
	Tetrachloroethene	20	18.5	ug/L	93	3		74	129	20
	Chlorobenzene	20	20.2	ug/L	101	1		82	118	20
	Ethyl Benzene	20	19.3	ug/L	97	1		79	121	20
	m/p-Xylenes	40	39.5	ug/L	99	0		80	121	20
	o-Xylene	20	20.0	ug/L	100	1		78	122	20
	Styrene	20	20.6	ug/L	103	2		78	123	20
	Bromoform	20	22.4	ug/L	112	6		66	130	20
	Isopropylbenzene	20	19.3	ug/L	97	0		72	131	20
	1,1,2,2-Tetrachloroethane	20	21.7	ug/L	109	4		71	121	20
	1,3-Dichlorobenzene	20	20.5	ug/L	103	1		80	119	20
	1,4-Dichlorobenzene	20	20.4	ug/L	102	0		79	118	20
	1,2-Dichlorobenzene	20	20.6	ug/L	103	3		80	119	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2263

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VN086945.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VN0611WBSD02	1,2-Dibromo-3-Chloropropane	20	20.8	ug/L	104	8		62	128	20
	1,2,4-Trichlorobenzene	20	18.7	ug/L	94	2		69	130	20
	1,2,3-Trichlorobenzene	20	18.1	ug/L	91	2		69	129	20

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VN0611WBL01

Lab Name: CHEMTECH

Contract: AECO15

Lab Code: CHEM Case No.: Q2263

SAS No.: Q2263 SDG NO.: Q2263

Lab File ID: VN086942.D

Lab Sample ID: VN0611WBL01

Date Analyzed: 06/11/2025

Time Analyzed: 12:28

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_N

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VN0611WBS02	VN0611WBS02	VN086944.D	06/11/2025
VN0611WBSD02	VN0611WBSD02	VN086945.D	06/11/2025
RW9-MW01D3-20250606	Q2263-01	VN086963.D	06/11/2025
RW9-MW01S-20250606	Q2263-02	VN086964.D	06/11/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q2263 SAS No.: Q2263 SDG NO.: Q2263
Lab File ID: VN086861.D BFB Injection Date: 06/06/2025
Instrument ID: MSVOA_N BFB Injection Time: 07:59
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	17.3
75	30.0 - 60.0% of mass 95	48.1
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.4
173	Less than 2.0% of mass 174	0.7 (1) 1
174	50.0 - 100.0% of mass 95	66.6
175	5.0 - 9.0% of mass 174	4.7 (7.1) 1
176	95.0 - 101.0% of mass 174	65.3 (98.1) 1
177	5.0 - 9.0% of mass 176	4.4 (6.8) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VN086862.D	06/06/2025	12:44
VSTDIC005	VSTDIC005	VN086863.D	06/06/2025	13:17
VSTDIC020	VSTDIC020	VN086864.D	06/06/2025	13:40
VSTDIC050	VSTDIC050	VN086865.D	06/06/2025	14:03
VSTDIC100	VSTDIC100	VN086866.D	06/06/2025	14:26
VSTDIC150	VSTDIC150	VN086867.D	06/06/2025	14:49

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q2263 SAS No.: Q2263 SDG NO.: Q2263
Lab File ID: VN086939.D BFB Injection Date: 06/11/2025
Instrument ID: MSVOA_N BFB Injection Time: 10:22
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	15.7
75	30.0 - 60.0% of mass 95	46.5
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7.1
173	Less than 2.0% of mass 174	0.7 (1) 1
174	50.0 - 100.0% of mass 95	71.1
175	5.0 - 9.0% of mass 174	5.4 (7.5) 1
176	95.0 - 101.0% of mass 174	68.8 (96.8) 1
177	5.0 - 9.0% of mass 176	4.8 (7) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VN086940.D	06/11/2025	11:32
VN0611WBL01	VN0611WBL01	VN086942.D	06/11/2025	12:28
VN0611WBS02	VN0611WBS02	VN086944.D	06/11/2025	13:27
VN0611WBSD02	VN0611WBSD02	VN086945.D	06/11/2025	14:01
RW9-MW01D3-20250606	Q2263-01	VN086963.D	06/11/2025	20:39
RW9-MW01S-20250606	Q2263-02	VN086964.D	06/11/2025	21:01
VSTDCCC050EC	VSTDCCC050	VN086965.D	06/11/2025	21:23

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q2263 SAS No.: Q2263 SDG NO.: Q2263
Lab File ID: VN086940.D Date Analyzed: 06/11/2025
Instrument ID: MSVOA_N Time Analyzed: 11:32
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	205098	8.23	357643	9.11	312173	11.87
UPPER LIMIT	410196	8.73	715286	9.606	624346	12.365
LOWER LIMIT	102549	7.73	178822	8.606	156087	11.365
EPA SAMPLE NO.						
RW9-MW01D3-20250606	255804	8.24	486937	9.11	433933	11.87
RW9-MW01S-20250606	212082	8.24	417259	9.11	370384	11.87
VN0611WBL01	328120	8.23	618651	9.11	543433	11.87
VN0611WBS02	206799	8.24	370524	9.11	321188	11.87
VN0611WBSD02	199904	8.23	354069	9.11	310598	11.87

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q2263 SAS No.: Q2263 SDG NO.: Q2263
 Lab File ID: VN086940.D Date Analyzed: 06/11/2025
 Instrument ID: MSVOA_N Time Analyzed: 11:32
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	152136	13.788				
UPPER LIMIT	304272	14.288				
LOWER LIMIT	76068	13.288				
EPA SAMPLE NO.						
RW9-MW01D3-20250606	207409	13.79				
RW9-MW01S-20250606	174250	13.79				
VN0611WBL01	257257	13.79				
VN0611WBS02	157403	13.79				
VN0611WBSD02	151691	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VN0611WBL01		SDG No.:	Q2263
Lab Sample ID:	VN0611WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086942.D	1		06/11/25 12:28	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VN0611WBL01		SDG No.:	Q2263
Lab Sample ID:	VN0611WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086942.D	1		06/11/25 12:28	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	48.3		81 - 118		97%	SPK: 50
1868-53-7	Dibromofluoromethane	49.3		80 - 119		99%	SPK: 50
2037-26-5	Toluene-d8	51.7		89 - 112		103%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.4		85 - 114		99%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	328000	8.23				
540-36-3	1,4-Difluorobenzene	619000	9.106				
3114-55-4	Chlorobenzene-d5	543000	11.865				
3855-82-1	1,4-Dichlorobenzene-d4	257000	13.788				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VN0611WBL01		SDG No.:	Q2263
Lab Sample ID:	VN0611WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086942.D	1		06/11/25 12:28	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VN0611WBS02		SDG No.:	Q2263
Lab Sample ID:	VN0611WBS02		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086944.D	1		06/11/25 13:27	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	19.7		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	16.1		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	18.9		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	16.6		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	19.4		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	19.2		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	19.6		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	19.2		0.23	0.75	1.00	ug/L
67-64-1	Acetone	99.0		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	17.8		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	19.6		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	19.0		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	19.0		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	18.5		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	19.4		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	17.6		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	90.4		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	19.2		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.5		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	19.9		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	19.5		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	19.0		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	16.9		0.16	0.50	1.00	ug/L
71-43-2	Benzene	19.3		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	19.5		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	19.9		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	19.6		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	19.6		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	98.0		0.68	2.50	5.00	ug/L
108-88-3	Toluene	19.5		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VN0611WBS02		SDG No.:	Q2263
Lab Sample ID:	VN0611WBS02		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086944.D	1		06/11/25 13:27	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	19.9		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	20.1		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	20.4		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	89.2		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	20.5		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	19.7		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	19.2		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	20.0		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	19.6		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	39.6		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	20.2		0.12	0.50	1.00	ug/L
100-42-5	Styrene	20.1		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	21.0		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	19.3		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	21.0		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	20.3		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	20.4		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	20.0		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	19.2		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	18.4		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	17.8		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	47.1		81 - 118		94%	SPK: 50
1868-53-7	Dibromofluoromethane	51.1		80 - 119		102%	SPK: 50
2037-26-5	Toluene-d8	48.7		89 - 112		97%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.0		85 - 114		100%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	207000	8.235				
540-36-3	1,4-Difluorobenzene	371000	9.106				
3114-55-4	Chlorobenzene-d5	321000	11.865				
3855-82-1	1,4-Dichlorobenzene-d4	157000	13.788				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VN0611WBS02		SDG No.:	Q2263
Lab Sample ID:	VN0611WBS02		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086944.D	1		06/11/25 13:27	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VN0611WBSD02		SDG No.:	Q2263
Lab Sample ID:	VN0611WBSD02		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086945.D	1		06/11/25 14:01	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	19.1		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	15.5		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	18.7		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	16.2		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	18.9		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	19.0		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	18.7		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	19.1		0.23	0.75	1.00	ug/L
67-64-1	Acetone	92.4		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	17.2		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	20.2		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	20.1		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	19.1		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	18.6		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	19.3		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	17.2		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	94.9		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	19.5		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.6		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	21.1		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	19.4		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	18.8		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	16.6		0.16	0.50	1.00	ug/L
71-43-2	Benzene	19.7		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	20.3		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	20.1		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	20.0		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	20.7		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	100		0.68	2.50	5.00	ug/L
108-88-3	Toluene	19.9		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VN0611WBSD02		SDG No.:	Q2263
Lab Sample ID:	VN0611WBSD02		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086945.D	1		06/11/25 14:01	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	20.7		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	20.7		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	20.9		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	93.8		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	21.2		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	20.7		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	18.5		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	20.2		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	19.3		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	39.5		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	20.0		0.12	0.50	1.00	ug/L
100-42-5	Styrene	20.6		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	22.4		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	19.3		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	21.7		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	20.5		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	20.4		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	20.6		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	20.8		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	18.7		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	18.1		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	47.9		81 - 118		96%	SPK: 50
1868-53-7	Dibromofluoromethane	52.3		80 - 119		105%	SPK: 50
2037-26-5	Toluene-d8	49.5		89 - 112		99%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.1		85 - 114		100%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	200000	8.229				
540-36-3	1,4-Difluorobenzene	354000	9.106				
3114-55-4	Chlorobenzene-d5	311000	11.865				
3855-82-1	1,4-Dichlorobenzene-d4	152000	13.788				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VN0611WBSD02		SDG No.:	Q2263
Lab Sample ID:	VN0611WBSD02		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086945.D	1		06/11/25 14:01	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q2263 SAS No.: Q2263 SDG No.: Q2263

Instrument ID: MSVOA_N Calibration Date(s): 06/06/2025 06/06/2025

Heated Purge: (Y/N) N Calibration Time(s): 12:44 14:49

GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:								
RRF001 = VN086862.D			RRF005 = VN086863.D			RRF020 = VN086864.D		
RRF050 = VN086865.D			RRF100 = VN086866.D			RRF150 = VN086867.D		
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
Dichlorodifluoromethane	0.467	0.444	0.539	0.501	0.535	0.506	0.499	7.5
Chloromethane	0.762	0.654	0.645	0.597	0.617	0.587	0.644	9.9
Vinyl Chloride	0.670	0.670	0.684	0.640	0.673	0.648	0.664	2.5
Bromomethane		0.375	0.380	0.357	0.379	0.368	0.372	2.6
Chloroethane	0.460	0.444	0.442	0.408	0.418	0.402	0.429	5.4
Trichlorofluoromethane	0.882	0.903	0.904	0.834	0.858	0.825	0.868	3.9
1,1,2-Trichlorotrifluoroethane	0.554	0.567	0.563	0.519	0.546	0.520	0.545	3.8
1,1-Dichloroethene	0.573	0.593	0.563	0.533	0.550	0.527	0.557	4.4
Acetone	0.426	0.366	0.366	0.322	0.334	0.316	0.355	11.5
Carbon Disulfide	1.718	1.622	1.542	1.426	1.496	1.433	1.539	7.4
Methyl tert-butyl Ether	2.120	2.038	2.051	1.933	2.021	1.926	2.015	3.7
Methyl Acetate	1.035	1.049	1.078	0.986	1.049	1.011	1.035	3.1
Methylene Chloride	0.822	0.688	0.643	0.605	0.629	0.601	0.665	12.5
trans-1,2-Dichloroethene	0.700	0.674	0.621	0.567	0.591	0.561	0.619	9.3
1,1-Dichloroethane	1.192	1.153	1.156	1.063	1.110	1.043	1.120	5.2
Cyclohexane		1.303	1.116	1.004	1.030	0.976	1.086	12.2
2-Butanone	0.604	0.598	0.604	0.551	0.573	0.533	0.577	5.2
Carbon Tetrachloride	0.453	0.449	0.434	0.409	0.435	0.421	0.433	3.9
cis-1,2-Dichloroethene	0.786	0.766	0.762	0.699	0.729	0.701	0.740	4.9
Bromochloromethane	0.579	0.564	0.616	0.466	0.517	0.560	0.550	9.5
Chloroform	1.235	1.152	1.145	1.061	1.085	1.030	1.118	6.7
1,1,1-Trichloroethane	1.029	0.995	0.969	0.895	0.925	0.893	0.951	5.9
Methylcyclohexane	0.633	0.645	0.588	0.570	0.603	0.589	0.605	4.8
Benzene	1.588	1.501	1.444	1.345	1.414	1.371	1.444	6.2
1,2-Dichloroethane	0.473	0.456	0.444	0.411	0.430	0.413	0.438	5.6
Trichloroethene	0.359	0.360	0.341	0.327	0.340	0.328	0.342	4.2
1,2-Dichloropropane	0.366	0.369	0.354	0.332	0.352	0.335	0.351	4.4
Bromodichloromethane	0.510	0.484	0.480	0.457	0.483	0.465	0.480	3.8
4-Methyl-2-Pentanone	0.505	0.549	0.576	0.538	0.562	0.528	0.543	4.6
Toluene	0.918	0.914	0.885	0.835	0.883	0.859	0.882	3.6

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name:	<u>CHEMTECH</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q2263</u>
Instrument ID:	<u>MSVOA_N</u>	SAS No.:	<u>Q2263</u>
Heated Purge: (Y/N)	<u>N</u>	SDG No.:	<u>Q2263</u>
GC Column:	<u>RXI-624</u>	Calibration Date(s):	<u>06/06/2025</u>
ID:	<u>0.25 (mm)</u>	Calibration Time(s):	<u>12:44</u>
			<u>14:49</u>

LAB FILE ID:								
RRF001 = VN086862.D			RRF005 = VN086863.D			RRF020 = VN086864.D		
RRF050 = VN086865.D			RRF100 = VN086866.D			RRF150 = VN086867.D		
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
t-1,3-Dichloropropene	0.571	0.526	0.524	0.522	0.548	0.530	0.537	3.6
cis-1,3-Dichloropropene	0.609	0.577	0.564	0.551	0.584	0.561	0.574	3.6
1,1,2-Trichloroethane	0.359	0.355	0.342	0.322	0.335	0.323	0.340	4.7
2-Hexanone	0.312	0.282	0.363	0.368	0.397	0.376	0.350	12.4
Dibromochloromethane	0.361	0.351	0.358	0.340	0.363	0.349	0.354	2.5
1,2-Dibromoethane	0.353	0.358	0.354	0.329	0.354	0.341	0.348	3.2
Tetrachloroethene	0.355	0.331	0.312	0.294	0.313	0.293	0.316	7.5
Chlorobenzene	1.233	1.135	1.107	1.023	1.089	1.030	1.103	7
Ethyl Benzene	1.989	1.975	1.907	1.796	1.913	1.816	1.900	4.2
m/p-Xylenes	0.730	0.751	0.741	0.701	0.736	0.703	0.727	2.8
o-Xylene	0.714	0.699	0.702	0.674	0.711	0.678	0.696	2.4
Styrene	1.177	1.193	1.226	1.162	1.229	1.164	1.192	2.5
Bromoform	0.217	0.266	0.276	0.265	0.286	0.267	0.263	9.1
Isopropylbenzene	3.864	3.749	3.649	3.426	3.621	3.546	3.643	4.2
1,1,2,2-Tetrachloroethane	1.292	1.299	1.273	1.178	1.205	1.157	1.234	5
1,3-Dichlorobenzene	1.763	1.709	1.657	1.554	1.612	1.566	1.644	5
1,4-Dichlorobenzene	1.820	1.786	1.657	1.572	1.642	1.576	1.676	6.3
1,2-Dichlorobenzene	1.675	1.651	1.596	1.500	1.557	1.496	1.579	4.8
1,2-Dibromo-3-Chloropropane	0.339	0.317	0.290	0.272	0.283	0.270	0.295	9.3
1,2,4-Trichlorobenzene	1.042	1.037	0.991	0.969	1.016	0.994	1.008	2.8
1,2,3-Trichlorobenzene	1.073	1.009	0.993	0.950	1.000	0.987	1.002	4
1,2-Dichloroethane-d4		0.732	0.707	0.500	0.656	0.751	0.669	15.1
Dibromofluoromethane		0.303	0.310	0.219	0.298	0.351	0.296	16.2
Toluene-d8		1.245	1.203	0.861	1.178	1.377	1.173	16.2
4-Bromofluorobenzene		0.441	0.446	0.325	0.446	0.521	0.436	16.2

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q2263 SAS No.: Q2263 SDG No.: Q2263

Instrument ID: MSVOA_N Calibration Date/Time: 06/11/2025 11:32

Lab File ID: VN086940.D Init. Calib. Date(s): 06/06/2025 06/06/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:44 14:49

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.499	0.538		7.82	20
Chloromethane	0.644	0.555	0.1	-13.82	20
Vinyl Chloride	0.664	0.701		5.57	20
Bromomethane	0.372	0.312		-16.13	20
Chloroethane	0.429	0.457		6.53	20
Trichlorofluoromethane	0.868	0.930		7.14	20
1,1,2-Trichlorotrifluoroethane	0.545	0.582		6.79	20
1,1-Dichloroethene	0.557	0.593		6.46	20
Acetone	0.355	0.384		8.17	20
Carbon Disulfide	1.539	1.505		-2.21	20
Methyl tert-butyl Ether	2.015	2.172		7.79	20
Methyl Acetate	1.035	1.085		4.83	20
Methylene Chloride	0.665	0.673		1.2	20
trans-1,2-Dichloroethene	0.619	0.633		2.26	20
1,1-Dichloroethane	1.120	1.178	0.1	5.18	20
Cyclohexane	1.086	1.017		-6.35	20
2-Butanone	0.577	0.568		-1.56	20
Carbon Tetrachloride	0.433	0.472		9.01	20
cis-1,2-Dichloroethene	0.740	0.793		7.16	20
Bromochloromethane	0.550	0.578		5.09	20
Chloroform	1.118	1.176		5.19	20
1,1,1-Trichloroethane	0.951	0.976		2.63	20
Methylcyclohexane	0.605	0.576		-4.79	20
Benzene	1.444	1.554		7.55	20
1,2-Dichloroethane	0.438	0.472		7.76	20
Trichloroethene	0.342	0.382		11.7	20
1,2-Dichloropropane	0.351	0.384		9.4	20
Bromodichloromethane	0.480	0.532		10.83	20
4-Methyl-2-Pentanone	0.543	0.590		8.66	20
Toluene	0.882	0.978		10.88	20
t-1,3-Dichloropropene	0.537	0.609		13.41	20
cis-1,3-Dichloropropene	0.574	0.661		15.16	20
1,1,2-Trichloroethane	0.340	0.383		12.65	20
2-Hexanone	0.350	0.395		12.86	20
Dibromochloromethane	0.354	0.414		16.95	20
1,2-Dibromoethane	0.348	0.385		10.63	20
Tetrachloroethene	0.316	0.337		6.65	20
Chlorobenzene	1.103	1.222	0.3	10.79	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q2263 SAS No.: Q2263 SDG No.: Q2263

Instrument ID: MSVOA_N Calibration Date/Time: 06/11/2025 11:32

Lab File ID: VN086940.D Init. Calib. Date(s): 06/06/2025 06/06/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:44 14:49

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.900	2.059		8.37	20
m/p-Xylenes	0.727	0.802		10.32	20
o-Xylene	0.696	0.782		12.36	20
Styrene	1.192	1.358		13.93	20
Bromoform	0.263	0.319	0.1	21.29	20
Isopropylbenzene	3.643	3.898		7	20
1,1,2,2-Tetrachloroethane	1.234	1.402	0.3	13.61	20
1,3-Dichlorobenzene	1.644	1.843		12.1	20
1,4-Dichlorobenzene	1.676	1.887		12.59	20
1,2-Dichlorobenzene	1.579	1.761		11.53	20
1,2-Dibromo-3-Chloropropane	0.295	0.311		5.42	20
1,2,4-Trichlorobenzene	1.008	1.052		4.36	20
1,2,3-Trichlorobenzene	1.002	0.995		-0.7	20
1,2-Dichloroethane-d4	0.669	0.622		-7.03	20
Dibromofluoromethane	0.296	0.309		4.39	20
Toluene-d8	1.173	1.181		0.68	20
4-Bromofluorobenzene	0.436	0.440		0.92	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q2263 SAS No.: Q2263 SDG No.: Q2263

Instrument ID: MSVOA_N Calibration Date/Time: 06/11/2025 21:23

Lab File ID: VN086965.D Init. Calib. Date(s): 06/06/2025 06/06/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:44 14:49

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.499	0.552		10.62	50
Chloromethane	0.644	0.581	0.1	-9.78	50
Vinyl Chloride	0.664	0.712		7.23	50
Bromomethane	0.372	0.246		-33.87	50
Chloroethane	0.429	0.477		11.19	50
Trichlorofluoromethane	0.868	0.969		11.64	50
1,1,2-Trichlorotrifluoroethane	0.545	0.587		7.71	50
1,1-Dichloroethene	0.557	0.599		7.54	50
Acetone	0.355	0.373		5.07	50
Carbon Disulfide	1.539	1.513		-1.75	50
Methyl tert-butyl Ether	2.015	2.289		13.6	50
Methyl Acetate	1.035	1.207		16.62	50
Methylene Chloride	0.665	0.707		6.32	50
trans-1,2-Dichloroethene	0.619	0.653		5.49	50
1,1-Dichloroethane	1.120	1.255	0.1	12.05	50
Cyclohexane	1.086	1.022		-5.89	50
2-Butanone	0.577	0.628		8.84	50
Carbon Tetrachloride	0.433	0.462		6.7	50
cis-1,2-Dichloroethene	0.740	0.824		11.35	50
Bromochloromethane	0.550	0.630		14.55	50
Chloroform	1.118	1.234		10.38	50
1,1,1-Trichloroethane	0.951	1.024		7.68	50
Methylcyclohexane	0.605	0.540		-10.74	50
Benzene	1.444	1.551		7.41	50
1,2-Dichloroethane	0.438	0.472		7.76	50
Trichloroethene	0.342	0.366		7.02	50
1,2-Dichloropropane	0.351	0.381		8.55	50
Bromodichloromethane	0.480	0.537		11.88	50
4-Methyl-2-Pentanone	0.543	0.610		12.34	50
Toluene	0.882	0.953		8.05	50
t-1,3-Dichloropropene	0.537	0.588		9.5	50
cis-1,3-Dichloropropene	0.574	0.623		8.54	50
1,1,2-Trichloroethane	0.340	0.379		11.47	50
2-Hexanone	0.350	0.405		15.71	50
Dibromochloromethane	0.354	0.407		14.97	50
1,2-Dibromoethane	0.348	0.391		12.36	50
Tetrachloroethene	0.316	0.320		1.27	50
Chlorobenzene	1.103	1.173	0.3	6.35	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q2263 SAS No.: Q2263 SDG No.: Q2263

Instrument ID: MSVOA_N Calibration Date/Time: 06/11/2025 21:23

Lab File ID: VN086965.D Init. Calib. Date(s): 06/06/2025 06/06/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:44 14:49

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.900	2.012		5.89	50
m/p-Xylenes	0.727	0.775		6.6	50
o-Xylene	0.696	0.766		10.06	50
Styrene	1.192	1.318		10.57	50
Bromoform	0.263	0.313	0.1	19.01	50
Isopropylbenzene	3.643	3.761		3.24	50
1,1,2,2-Tetrachloroethane	1.234	1.414	0.3	14.59	50
1,3-Dichlorobenzene	1.644	1.784		8.52	50
1,4-Dichlorobenzene	1.676	1.781		6.26	50
1,2-Dichlorobenzene	1.579	1.721		8.99	50
1,2-Dibromo-3-Chloropropane	0.295	0.323		9.49	50
1,2,4-Trichlorobenzene	1.008	0.998		-0.99	50
1,2,3-Trichlorobenzene	1.002	0.977		-2.49	50
1,2-Dichloroethane-d4	0.669	0.697		4.18	50
Dibromofluoromethane	0.296	0.314		6.08	50
Toluene-d8	1.173	1.171		-0.17	50
4-Bromofluorobenzene	0.436	0.449		2.98	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.